

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ112918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 30 00:42:36 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ035073.D 1600 =PQ035075.D 800 =PQ035074.D
 200 =PQ035072.D 100 =PQ035071.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.659	3.887	4.003	3.576	2.913	3.608	E6	11.77
2) SA	Decachlorobiphenyl	3.623	3.317	2.887	3.675	3.904	3.481	E6	11.27
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	1.628	1.651	1.668	1.650	1.529	1.625	E5	3.43
27) L6	AR-1254-2	2.577	2.623	2.616	2.634	2.349	2.560	E5	4.68
28) L6	AR-1254-3	2.789	2.923	2.844	2.806	2.598	2.792	E5	4.30
29) L6	AR-1254-4	2.024	2.081	2.006	2.060	1.883	2.011	E5	3.85
30) L6	AR-1254-5	2.236	2.288	2.159	2.249	1.969	2.180	E5	5.82
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ035073.D 1600 =PQ035075.D 800 =PQ035074.D
 200 =PQ035072.D 100 =PQ035071.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.111	2.228	2.344	2.056	1.662	2.080	E6	12.45
2) SA	Decachlorobiphenyl	2.289	2.103	1.883	2.360	2.429	2.213	E6	9.99
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ112918CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 30 00:42:36 2018
 Response Via : Initial Calibration

Calibration Files

400 =PQ035073.D 1600 =PQ035075.D 800 =PQ035074.D
 200 =PQ035072.D 100 =PQ035071.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.449	1.465	1.501	1.466	1.301	1.436 E5	5.44
27) L6	AR-1254-2	1.251	1.253	1.282	1.277	1.129	1.239 E5	5.06
28) L6	AR-1254-3	2.120	2.181	2.168	2.153	1.946	2.114 E5	4.55
29) L6	AR-1254-4	1.333	1.371	1.350	1.373	1.205	1.326 E5	5.25
30) L6	AR-1254-5	1.846	1.863	1.799	1.893	1.731	1.826 E5	3.47
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range