

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112918\
 Data File : PQ035075.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Nov 2018 23:28
 Operator : SM\SJ
 Sample : AR1254ICC1600
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 30 00:33:13 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112918CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 30 00:29:24 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.455	3.753	311.0E6	178.2E6	82.420	82.148
2) SA Decachlor...	10.193	8.751	530.7E6	336.4E6	152.943	153.203
Target Compounds						
26) L6 AR-1254-1	6.477	5.635	264.1E6	234.4E6	1611.277	1608.962
27) L6 AR-1254-2	6.694	5.781	419.6E6	200.6E6	1614.158	1601.342
28) L6 AR-1254-3	7.064	6.184	467.7E6	349.0E6	1637.520	1622.888
29) L6 AR-1254-4	7.349	6.411	333.0E6	219.3E6	1622.360	1622.673
30) L6 AR-1254-5	7.768	6.827	366.0E6	298.1E6	1618.228	1607.486

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ112918\
Data File : PQ035075.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Nov 2018 23:28
Operator : SM\SJ
Sample : AR1254ICC1600
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 30 00:33:13 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ112918CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 30 00:29:24 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

