

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ121318.M
 Title : GC EXTRACTABLES
 Last Update : Wed Dec 12 16:27:59 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ035375.D 1000 =PQ035373.D 750 =PQ035374.D
 250 =PQ035376.D 50 =PQ035377.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.293	4.307	4.030	3.667	3.265	3.913	E6	11.38
2) SA	Decachlorobiphenyl	2.744	2.592	2.942	2.901	3.560	2.948	E6	12.53
3) L1	AR-1016-1	1.335	1.304	1.247	1.236	1.276	1.279	E5	3.18
4) L1	AR-1016-2	2.004	1.936	1.890	1.835	1.925	1.918	E5	3.23
5) L1	AR-1016-3	1.178	1.146	1.095	1.090	1.213	1.144	E5	4.64
6) L1	AR-1016-4	0.957	0.945	0.906	0.920	1.034	0.952	E5	5.24
7) L1	AR-1016-5	9.420	9.140	8.884	8.869	9.205	9.104	E4	2.55
8) L2	AR-1221-1	3.416					3.416	E4	0.00
9) L2	AR-1221-2	2.548					2.548	E4	0.00
10) L2	AR-1221-3	8.322					8.322	E4	0.00
11) L3	AR-1232-1	7.272					7.272	E4	0.00
12) L3	AR-1232-2	3.854					3.854	E4	0.00
13) L3	AR-1232-3	8.177					8.177	E4	0.00
14) L3	AR-1232-4	4.004					4.004	E4	0.00
15) L3	AR-1232-5	3.167					3.167	E4	0.00
16) L4	AR-1242-1	9.648					9.648	E4	0.00
17) L4	AR-1242-2	1.451					1.451	E5	0.00
18) L4	AR-1242-3	8.652					8.652	E4	0.00
19) L4	AR-1242-4	6.918					6.918	E4	0.00
20) L4	AR-1242-5	7.845					7.845	E4	0.00
21) L5	AR-1248-1	7.991					7.991	E4	0.00
22) L5	AR-1248-2	9.478					9.478	E4	0.00
23) L5	AR-1248-3	1.045					1.045	E5	0.00
24) L5	AR-1248-4	1.231					1.231	E5	0.00
25) L5	AR-1248-5	1.170					1.170	E5	0.00
26) L6	AR-1254-1	7.295					7.295	E4	0.00
27) L6	AR-1254-2	2.366					2.366	E5	0.00
28) L6	AR-1254-3	2.568					2.568	E5	0.00
29) L6	AR-1254-4	1.813					1.813	E5	0.00
30) L6	AR-1254-5	1.962					1.962	E5	0.00
31) L7	AR-1260-1	1.823	1.727	1.809	1.844	2.016	1.844	E5	5.74
32) L7	AR-1260-2	2.126	2.007	2.160	2.164	2.407	2.173	E5	6.71
33) L7	AR-1260-3	1.279	1.204	1.301	1.327	1.588	1.340	E5	10.90
34) L7	AR-1260-4	1.567	1.477	1.646	1.721	1.830	1.648	E5	8.27
35) L7	AR-1260-5	2.818	2.702	3.096	2.997	3.294	2.982	E5	7.79
36) L8	AR-1262-1	2.191					2.191	E5	0.00
37) L8	AR-1262-2	3.767					3.767	E5	0.00
38) L8	AR-1262-3	2.404					2.404	E5	0.00
39) L8	AR-1262-4	1.807					1.807	E5	0.00
40) L8	AR-1262-5	1.139					1.139	E5	0.00
41) L9	AR-1268-1	4.825					4.825	E5	0.00
42) L9	AR-1268-2	4.315					4.315	E5	0.00
43) L9	AR-1268-3	3.779					3.779	E5	0.00
44) L9	AR-1268-4	1.371					1.371	E5	0.00
45) L9	AR-1268-5	1.158					1.158	E6	0.00

Signal #2 Calibration Files

500 =PQ035375.D 1000 =PQ035373.D 750 =PQ035374.D
 250 =PQ035376.D 50 =PQ035377.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.443	2.423	2.285	2.168	1.854	2.234	E6	10.75
2) SA	Decachlorobiphenyl	1.665	1.549	1.736	1.738	2.149	1.767	E6	12.84
3) L1	AR-1016-1	8.083	7.857	7.598	7.302	7.712	7.710	E4	3.78
4) L1	AR-1016-2	1.181	1.163	1.127	1.096	1.090	1.132	E5	3.56
5) L1	AR-1016-3	6.049	5.913	5.770	5.751	5.717	5.840	E4	2.38

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ121318.M
 Title : GC EXTRACTABLES
 Last Update : Wed Dec 12 16:27:59 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ035375.D 1000 =PQ035373.D 750 =PQ035374.D
 250 =PQ035376.D 50 =PQ035377.D

Compound			500	1000	750	250	50	Avg		%RSD
6)	L1	AR-1016-4	4.874	4.642	4.576	4.818	4.599	4.702	E4	2.88
7)	L1	AR-1016-5	6.436	6.114	5.996	6.018	6.130	6.139	E4	2.87
8)	L2	AR-1221-1	2.040					2.040	E4	0.00
9)	L2	AR-1221-2	1.478					1.478	E4	0.00
10)	L2	AR-1221-3	4.858					4.858	E4	0.00
11)	L3	AR-1232-1	4.403					4.403	E4	0.00
12)	L3	AR-1232-2	4.906					4.906	E4	0.00
13)	L3	AR-1232-3	2.555					2.555	E4	0.00
14)	L3	AR-1232-4	2.282					2.282	E4	0.00
15)	L3	AR-1232-5	2.498					2.498	E4	0.00
16)	L4	AR-1242-1	5.917					5.917	E4	0.00
17)	L4	AR-1242-2	8.611					8.611	E4	0.00
18)	L4	AR-1242-3	4.504					4.504	E4	0.00
19)	L4	AR-1242-4	4.428					4.428	E4	0.00
20)	L4	AR-1242-5	5.519					5.519	E4	0.00
21)	L5	AR-1248-1	3.945					3.945	E4	0.00
22)	L5	AR-1248-2	5.427					5.427	E4	0.00
23)	L5	AR-1248-3	5.580					5.580	E4	0.00
24)	L5	AR-1248-4	6.923					6.923	E4	0.00
25)	L5	AR-1248-5	7.048					7.048	E4	0.00
26)	L6	AR-1254-1	1.334					1.334	E5	0.00
27)	L6	AR-1254-2	1.155					1.155	E5	0.00
28)	L6	AR-1254-3	1.931					1.931	E5	0.00
29)	L6	AR-1254-4	1.203					1.203	E5	0.00
30)	L6	AR-1254-5	1.629					1.629	E5	0.00
31)	L7	AR-1260-1	1.228	1.191	1.225	1.226	1.419	1.258	E5	7.27
32)	L7	AR-1260-2	1.452	1.406	1.485	1.473	1.735	1.510	E5	8.54
33)	L7	AR-1260-3	1.370	1.342	1.404	1.373	1.499	1.397	E5	4.34
34)	L7	AR-1260-4	0.888	0.851	0.930	0.924	1.049	0.929	E5	8.00
35)	L7	AR-1260-5	2.133	2.069	2.322	2.171	2.525	2.244	E5	8.13
36)	L8	AR-1262-1	1.535					1.535	E5	0.00
37)	L8	AR-1262-2	2.669					2.669	E5	0.00
38)	L8	AR-1262-3	1.038					1.038	E5	0.00
39)	L8	AR-1262-4	1.852					1.852	E5	0.00
40)	L8	AR-1262-5	7.423					7.423	E4	0.00
41)	L9	AR-1268-1	3.335					3.335	E5	0.00
42)	L9	AR-1268-2	2.964					2.964	E5	0.00
43)	L9	AR-1268-3	2.444					2.444	E5	0.00
44)	L9	AR-1268-4	9.329					9.329	E4	0.00
45)	L9	AR-1268-5	7.304					7.304	E5	0.00

(#) = Out of Range