

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121219\
 Data File : PQ045806.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2019 13:51
 Operator : SM\AJ
 Sample : K6208-01DL 2X
 Misc :
 ALS Vial : 141 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 CONCRETE-COMP-1DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 15:04:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 11 11:57:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.248	4.370	23506843	23601391	14.635	13.814
2)	SA Decachlor...	11.417	9.816	93920815	71897613	12.646	12.371
Target Compounds							
26)	L6 AR-1254-1	7.474	6.535	50061244	72064257	386.098	426.823
27)	L6 AR-1254-2	7.703	6.692	134.7E6	110.6E6	615.083	700.669
28)	L6 AR-1254-3	8.086	7.122	199.7E6	229.4E6	807.424	840.750
29)	L6 AR-1254-4	8.380	7.361	256.9E6	220.8E6	1250.451	1289.352
30)	L6 AR-1254-5	8.811	7.795	364.1E6	450.9E6	1538.645	1623.804

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121219\
Data File : PQ045806.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2019 13:51
Operator : SM\AJ
Sample : K6208-01DL 2X
Misc :
ALS Vial : 141 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
CONCRETE-COMP-1DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 15:04:30 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121119.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 11 11:57:47 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

