

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ121719.M
 Title : GC EXTRACTABLES
 Last Update : Wed Dec 18 07:45:22 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ045842.D 1000 =PQ045840.D 750 =PQ045841.D
 250 =PQ045843.D 50 =PQ045844.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	6.035	5.114	5.588	6.265	7.351	6.071 E6	13.85
2) SA	Decachlorobiphenyl	1.166	1.109	1.134	1.197	1.330	1.187 E7	7.30
3) L1	AR-1016-1	2.073	1.760	1.966	2.219	2.552	2.114 E5	14.03
4) L1	AR-1016-2	3.069	2.656	2.919	3.268	3.664	3.115 E5	12.19
5) L1	AR-1016-3	1.924	1.665	1.832	2.038	2.234	1.939 E5	11.03
6) L1	AR-1016-4	1.623	1.415	1.556	1.731	1.986	1.662 E5	12.87
7) L1	AR-1016-5	1.604	1.430	1.583	1.753	1.892	1.652 E5	10.66
8) L2	AR-1221-1	6.246					6.246 E4	0.00
9) L2	AR-1221-2	4.551					4.551 E4	0.00
10) L2	AR-1221-3	1.346					1.346 E5	0.00
11) L3	AR-1232-1	2.315					2.315 E5	0.00
12) L3	AR-1232-2	1.246					1.246 E5	0.00
13) L3	AR-1232-3	2.533					2.533 E5	0.00
14) L3	AR-1232-4	1.345					1.345 E5	0.00
15) L3	AR-1232-5	9.999					9.999 E4	0.00
16) L4	AR-1242-1	1.535					1.535 E5	0.00
17) L4	AR-1242-2	2.364					2.364 E5	0.00
18) L4	AR-1242-3	1.475					1.475 E5	0.00
19) L4	AR-1242-4	1.261					1.261 E5	0.00
20) L4	AR-1242-5	1.610					1.610 E5	0.00
21) L5	AR-1248-1	1.231					1.231 E5	0.00
22) L5	AR-1248-2	1.774					1.774 E5	0.00
23) L5	AR-1248-3	2.065					2.065 E5	0.00
24) L5	AR-1248-4	2.697					2.697 E5	0.00
25) L5	AR-1248-5	2.715					2.715 E5	0.00
26) L6	AR-1254-1	3.186					3.186 E5	0.00
27) L6	AR-1254-2	5.287					5.287 E5	0.00
28) L6	AR-1254-3	6.129					6.129 E5	0.00
29) L6	AR-1254-4	5.262					5.262 E5	0.00
30) L6	AR-1254-5	5.721					5.721 E5	0.00
31) L7	AR-1260-1	3.946	3.523	3.804	4.253	5.029	4.111 E5	14.03
32) L7	AR-1260-2	5.377	4.850	5.221	5.706	6.123	5.455 E5	8.87
33) L7	AR-1260-3	4.803	4.343	4.628	4.930	5.552	4.851 E5	9.26
34) L7	AR-1260-4	4.833	4.446	4.687	4.895	5.348	4.842 E5	6.84
35) L7	AR-1260-5	1.135	1.074	1.119	1.116	1.266	1.142 E6	6.36
36) L8	AR-1262-1	8.383					8.383 E5	0.00
37) L8	AR-1262-2	1.565					1.565 E6	0.00
38) L8	AR-1262-3	1.047					1.047 E6	0.00
39) L8	AR-1262-4	8.597					8.597 E5	0.00
40) L8	AR-1262-5	5.650					5.650 E5	0.00
41) L9	AR-1268-1	1.887					1.887 E6	0.00
42) L9	AR-1268-2	1.786					1.786 E6	0.00
43) L9	AR-1268-3	1.534					1.534 E6	0.00
44) L9	AR-1268-4	6.234					6.234 E5	0.00
45) L9	AR-1268-5	4.900					4.900 E6	0.00

Signal #2 Calibration Files

500 =PQ045842.D 1000 =PQ045840.D 750 =PQ045841.D
 250 =PQ045843.D 50 =PQ045844.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	5.256	4.705	5.105	5.071	5.305	5.089 E6	4.64
2) SA	Decachlorobiphenyl	7.702	7.487	7.690	7.718	9.266	7.973 E6	9.15
3) L1	AR-1016-1	1.924	1.709	1.863	1.954	2.245	1.939 E5	10.08
4) L1	AR-1016-2	2.965	2.671	2.906	3.033	3.339	2.983 E5	8.09
5) L1	AR-1016-3	1.522	1.356	1.479	1.548	1.734	1.528 E5	8.95

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ121719.M
 Title : GC EXTRACTABLES
 Last Update : Wed Dec 18 07:45:22 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ045842.D 1000 =PQ045840.D 750 =PQ045841.D
 250 =PQ045843.D 50 =PQ045844.D

	Compound	500	1000	750	250	50	Avg	%RSD
6)	L1 AR-1016-4	1.304	1.137	1.246	1.376	1.612	1.335 E5	13.31
7)	L1 AR-1016-5	1.678	1.485	1.640	1.779	2.092	1.735 E5	13.03
8)	L2 AR-1221-1	5.108					5.108 E4	0.00
9)	L2 AR-1221-2	3.899					3.899 E4	0.00
10)	L2 AR-1221-3	1.226					1.226 E5	0.00
11)	L3 AR-1232-1	2.083					2.083 E5	0.00
12)	L3 AR-1232-2	2.417					2.417 E5	0.00
13)	L3 AR-1232-3	1.210					1.210 E5	0.00
14)	L3 AR-1232-4	1.145					1.145 E5	0.00
15)	L3 AR-1232-5	1.382					1.382 E5	0.00
16)	L4 AR-1242-1	1.440					1.440 E5	0.00
17)	L4 AR-1242-2	2.211					2.211 E5	0.00
18)	L4 AR-1242-3	1.145					1.145 E5	0.00
19)	L4 AR-1242-4	1.197					1.197 E5	0.00
20)	L4 AR-1242-5	1.674					1.674 E5	0.00
21)	L5 AR-1248-1	1.116					1.116 E5	0.00
22)	L5 AR-1248-2	1.644					1.644 E5	0.00
23)	L5 AR-1248-3	1.707					1.707 E5	0.00
24)	L5 AR-1248-4	2.109					2.109 E5	0.00
25)	L5 AR-1248-5	2.186					2.186 E5	0.00
26)	L6 AR-1254-1	4.067					4.067 E5	0.00
27)	L6 AR-1254-2	3.635					3.635 E5	0.00
28)	L6 AR-1254-3	6.384					6.384 E5	0.00
29)	L6 AR-1254-4	4.000					4.000 E5	0.00
30)	L6 AR-1254-5	6.038					6.038 E5	0.00
31)	L7 AR-1260-1	3.544	3.260	3.498	3.717	4.261	3.656 E5	10.27
32)	L7 AR-1260-2	4.382	4.023	4.325	4.565	5.291	4.517 E5	10.50
33)	L7 AR-1260-3	4.185	3.915	4.162	4.290	4.743	4.259 E5	7.12
34)	L7 AR-1260-4	3.738	3.537	3.733	3.814	4.243	3.813 E5	6.86
35)	L7 AR-1260-5	9.481	9.152	9.586	9.366	9.592	9.435 E5	1.94
36)	L8 AR-1262-1	3.364					3.364 E5	0.00
37)	L8 AR-1262-2	1.296					1.296 E6	0.00
38)	L8 AR-1262-3	4.903					4.903 E5	0.00
39)	L8 AR-1262-4	9.423					9.423 E5	0.00
40)	L8 AR-1262-5	3.968					3.968 E5	0.00
41)	L9 AR-1268-1	1.414					1.414 E6	0.00
42)	L9 AR-1268-2	1.327					1.327 E6	0.00
43)	L9 AR-1268-3	1.081					1.081 E6	0.00
44)	L9 AR-1268-4	4.351					4.351 E5	0.00
45)	L9 AR-1268-5	3.386					3.386 E6	0.00

(#) = Out of Range