

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121818\
 Data File : PQ035705.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2018 00:21
 Operator : SM\SJ
 Sample : J6389-11
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 P001-SS014-1218-01

Manual Integrations
 APPROVED

Sohil
 12/19/2018 10:57:44 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 19 06:08:14 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 12 22:07:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.428	3.742	106.9E6	57775373	27.322	25.857
2) SA Decachlor...	10.124	8.719	67426691	32074315	22.875m	18.147
Target Compounds						
31) L7 AR-1260-1	7.188	6.289	8716.1E6	5044.2E6	47266.882	40101.216
32) L7 AR-1260-2	7.445	6.477	9567.7E6	5653.7E6	44032.945	37438.389
33) L7 AR-1260-3	7.802	6.629	6447.4E6	5464.1E6	48128.837	39103.239
34) L7 AR-1260-4	8.031	7.097	8098.8E6	4254.6E6	49139.735	45819.450
35) L7 AR-1260-5	8.351	7.340	14254.3E6	8671.2E6	47806.581	38641.980

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121818\
Data File : PQ035705.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2018 00:21
Operator : SM\SJ
Sample : J6389-11
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
P001-SS014-1218-01

Manual Integrations
APPROVED

Sohil
12/19/2018 10:57:44 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 19 06:08:14 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121318.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 12 22:07:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

