

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121818\
 Data File : PQ035714.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2018 02:40
 Operator : SM\SJ
 Sample : J6388-10
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 P001-SS010-0006-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 19 06:10:32 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Dec 12 22:07:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.427	3.741	80925712	45896021	20.683	20.541
2) SA Decachlor...	10.123	8.718	40886319	21152031	13.871	11.967
Target Compounds						
31) L7 AR-1260-1	7.186	6.289	1344.1E6	902.8E6	7288.868	7177.227
32) L7 AR-1260-2	7.443	6.476	1510.3E6	1024.1E6	6950.649	6781.272
33) L7 AR-1260-3	7.801	6.629	949.7E6	957.6E6	7089.013	6853.314
34) L7 AR-1260-4	8.030	7.096	1174.5E6	685.5E6	7126.165	7382.663
35) L7 AR-1260-5	8.349	7.339	2203.7E6	1570.1E6	7390.789	6997.130

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121818\
Data File : PQ035714.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2018 02:40
Operator : SM\SJ
Sample : J6388-10
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
P001-SS010-0006-01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 19 06:10:32 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121318.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Dec 12 22:07:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

