

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ121919.M
 Title : GC EXTRACTABLES
 Last Update : Fri Dec 20 08:42:12 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ045910.D 1000 =PQ045908.D 750 =PQ045909.D
 250 =PQ045911.D 50 =PQ045912.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.944	3.875	3.835	3.644	4.198	3.899 E6	5.15
2) SA	Decachlorobiphenyl	1.006	1.006	0.988	0.903	1.199	1.020 E7	10.66
3) L1	AR-1016-1	1.283	1.220	1.221	1.214	1.414	1.270 E5	6.69
4) L1	AR-1016-2	1.922	1.829	1.885	1.842	2.084	1.912 E5	5.36
5) L1	AR-1016-3	1.213	1.161	1.197	1.153	1.302	1.205 E5	4.94
6) L1	AR-1016-4	0.993	0.948	0.988	0.940	1.009	0.976 E5	3.07
7) L1	AR-1016-5	1.016	0.969	0.993	0.960	1.145	1.016 E5	7.37
8) L2	AR-1221-1	3.701					3.701 E4	0.00
9) L2	AR-1221-2	2.835					2.835 E4	0.00
10) L2	AR-1221-3	8.498					8.498 E4	0.00
11) L3	AR-1232-1	1.448					1.448 E5	0.00
12) L3	AR-1232-2	7.160					7.160 E4	0.00
13) L3	AR-1232-3	1.435					1.435 E5	0.00
14) L3	AR-1232-4	7.317					7.317 E4	0.00
15) L3	AR-1232-5	5.570					5.570 E4	0.00
16) L4	AR-1242-1	9.723					9.723 E4	0.00
17) L4	AR-1242-2	1.450					1.450 E5	0.00
18) L4	AR-1242-3	9.359					9.359 E4	0.00
19) L4	AR-1242-4	7.569					7.569 E4	0.00
20) L4	AR-1242-5	9.209					9.209 E4	0.00
21) L5	AR-1248-1	8.421					8.421 E4	0.00
22) L5	AR-1248-2	1.201					1.201 E5	0.00
23) L5	AR-1248-3	1.365					1.365 E5	0.00
24) L5	AR-1248-4	1.684					1.684 E5	0.00
25) L5	AR-1248-5	1.614					1.614 E5	0.00
26) L6	AR-1254-1	1.498					1.498 E5	0.00
27) L6	AR-1254-2	2.494					2.494 E5	0.00
28) L6	AR-1254-3	2.962					2.962 E5	0.00
29) L6	AR-1254-4	2.650					2.650 E5	0.00
30) L6	AR-1254-5	3.150					3.150 E5	0.00
31) L7	AR-1260-1	2.461	2.407	2.430	2.330	2.744	2.474 E5	6.40
32) L7	AR-1260-2	3.491	3.453	3.423	3.173	3.454	3.399 E5	3.79
33) L7	AR-1260-3	3.276	3.205	3.239	2.980	3.401	3.220 E5	4.77
34) L7	AR-1260-4	3.445	3.397	3.396	3.176	3.962	3.475 E5	8.39
35) L7	AR-1260-5	8.677	8.894	8.658	7.640	8.747	8.523 E5	5.89
36) L8	AR-1262-1	4.784					4.784 E5	0.00
37) L8	AR-1262-2	1.009					1.009 E6	0.00
38) L8	AR-1262-3	7.345					7.345 E5	0.00
39) L8	AR-1262-4	6.083					6.083 E5	0.00
40) L8	AR-1262-5	4.639					4.639 E5	0.00
41) L9	AR-1268-1	1.336					1.336 E6	0.00
42) L9	AR-1268-2	1.262					1.262 E6	0.00
43) L9	AR-1268-3	1.067					1.067 E6	0.00
44) L9	AR-1268-4	5.552					5.552 E5	0.00
45) L9	AR-1268-5	3.941					3.941 E6	0.00

Signal #2 Calibration Files

500 =PQ045910.D 1000 =PQ045908.D 750 =PQ045909.D
 250 =PQ045911.D 50 =PQ045912.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.997	3.132	3.062	2.622	2.620	2.887 E6	8.56
2) SA	Decachlorobiphenyl	7.096	7.178	6.994	6.053	7.764	7.017 E6	8.78
3) L1	AR-1016-1	1.206	1.228	1.199	1.118	1.198	1.190 E5	3.50
4) L1	AR-1016-2	1.872	1.894	1.903	1.678	1.787	1.827 E5	5.20
5) L1	AR-1016-3	9.379	9.543	9.534	8.421	8.479	9.071 E4	6.30
6) L1	AR-1016-4	8.261	8.055	8.131	7.680	7.779	7.981 E4	3.06

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ121919.M
 Title : GC EXTRACTABLES
 Last Update : Fri Dec 20 08:42:12 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ045910.D 1000 =PQ045908.D 750 =PQ045909.D
 250 =PQ045911.D 50 =PQ045912.D

Compound			500	1000	750	250	50	Avg	%RSD

7)	L1	AR-1016-5	1.092	1.070	1.080	1.008	1.137	1.077 E5	4.33
8)	L2	AR-1221-1	2.829					2.829 E4	0.00
9)	L2	AR-1221-2	2.205					2.205 E4	0.00
10)	L2	AR-1221-3	6.927					6.927 E4	0.00
11)	L3	AR-1232-1	1.191					1.191 E5	0.00
12)	L3	AR-1232-2	1.363					1.363 E5	0.00
13)	L3	AR-1232-3	6.716					6.716 E4	0.00
14)	L3	AR-1232-4	6.691					6.691 E4	0.00
15)	L3	AR-1232-5	7.284					7.284 E4	0.00
16)	L4	AR-1242-1	9.013					9.013 E4	0.00
17)	L4	AR-1242-2	1.369					1.369 E5	0.00
18)	L4	AR-1242-3	6.965					6.965 E4	0.00
19)	L4	AR-1242-4	7.666					7.666 E4	0.00
20)	L4	AR-1242-5	1.022					1.022 E5	0.00
21)	L5	AR-1248-1	7.734					7.734 E4	0.00
22)	L5	AR-1248-2	1.152					1.152 E5	0.00
23)	L5	AR-1248-3	1.206					1.206 E5	0.00
24)	L5	AR-1248-4	1.474					1.474 E5	0.00
25)	L5	AR-1248-5	1.540					1.540 E5	0.00
26)	L6	AR-1254-1	2.131					2.131 E5	0.00
27)	L6	AR-1254-2	1.985					1.985 E5	0.00
28)	L6	AR-1254-3	3.497					3.497 E5	0.00
29)	L6	AR-1254-4	2.294					2.294 E5	0.00
30)	L6	AR-1254-5	3.519					3.519 E5	0.00
31)	L7	AR-1260-1	2.484	2.486	2.461	2.256	2.552	2.448 E5	4.60
32)	L7	AR-1260-2	3.112	3.163	3.043	2.834	3.215	3.074 E5	4.82
33)	L7	AR-1260-3	2.927	2.998	2.928	2.580	2.857	2.858 E5	5.71
34)	L7	AR-1260-4	2.717	2.775	2.696	2.377	2.776	2.668 E5	6.24
35)	L7	AR-1260-5	7.328	7.604	7.325	6.126	6.797	7.036 E5	8.34
36)	L8	AR-1262-1	1.908					1.908 E5	0.00
37)	L8	AR-1262-2	8.232					8.232 E5	0.00
38)	L8	AR-1262-3	3.193					3.193 E5	0.00
39)	L8	AR-1262-4	6.381					6.381 E5	0.00
40)	L8	AR-1262-5	3.104					3.104 E5	0.00
41)	L9	AR-1268-1	9.514					9.514 E5	0.00
42)	L9	AR-1268-2	9.153					9.153 E5	0.00
43)	L9	AR-1268-3	7.292					7.292 E5	0.00
44)	L9	AR-1268-4	3.344					3.344 E5	0.00
45)	L9	AR-1268-5	2.825					2.825 E6	0.00

(#) = Out of Range