

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121919\
 Data File : PQ045919.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2019 18:42
 Operator : SM\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 08:38:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 20 08:38:22 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.233	4.358	199.5E6	152.5E6	50.000	50.000
2) SA Decachlor...	11.400	9.803	933.7E6	661.2E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	9.787	8.627	667.9E6	475.7E6	500.000	500.000
42) L9 AR-1268-2	9.889	8.694	631.1E6	457.6E6	500.000	500.000
43) L9 AR-1268-3	10.134	8.907	533.6E6	364.6E6	500.000	500.000
44) L9 AR-1268-4	10.592	9.210	277.6E6	167.2E6	500.000	500.000
45) L9 AR-1268-5	11.036	9.525	1970.5E6	1412.5E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ121919\
Data File : PQ045919.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2019 18:42
Operator : SM\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 08:38:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ121919.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 20 08:38:22 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

