

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR010220CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jan 03 04:50:43 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR044229.D 1600 =PR044231.D 800 =PR044230.D
 200 =PR044228.D 100 =PR044227.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.360	2.525	3.092	3.661	3.560	3.239	E6	14.03
2) SA	Decachlorobiphenyl	5.304	4.469	4.824	5.653	5.820	5.214	E6	10.83
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	1.585	1.077	1.338	1.781	1.782	1.513	E5	20.12
27) L6	AR-1254-2	2.569	1.777	2.174	2.864	2.849	2.447	E5	19.10
28) L6	AR-1254-3	2.857	2.073	2.463	3.099	3.086	2.716	E5	16.28
29) L6	AR-1254-4	2.263	1.638	1.937	2.444	2.446	2.146	E5	16.40
30) L6	AR-1254-5	2.527	1.889	2.187	2.700	2.709	2.403	E5	14.84
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR044229.D 1600 =PR044231.D 800 =PR044230.D
 200 =PR044228.D 100 =PR044227.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	8.620	6.319	7.779	9.538	9.452	8.342	E6	16.03
2) SA	Decachlorobiphenyl	1.110	0.949	1.021	1.156	1.161	1.079	E7	8.52
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR010220CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jan 03 04:50:43 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR044229.D 1600 =PR044231.D 800 =PR044230.D
 200 =PR044228.D 100 =PR044227.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	4.351	2.981	3.674	4.845	4.779	4.126 E5	19.19
27) L6	AR-1254-2	4.054	2.784	3.409	4.530	4.494	3.854 E5	19.46
28) L6	AR-1254-3	7.009	5.024	6.016	7.691	7.744	6.697 E5	17.42
29) L6	AR-1254-4	5.440	3.931	4.659	5.924	5.984	5.188 E5	16.97
30) L6	AR-1254-5	6.774	5.194	6.111	7.494	6.878	6.490 E5	13.49
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range