

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR010419CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jan 05 01:14:53 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035224.D 1600 =PR035226.D 800 =PR035225.D
 200 =PR035223.D 100 =PR035222.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	0.926	1.102	1.065	0.934	0.967	0.999	E6	7.98
2)	SA	Decachlorobiphenyl	2.183	2.081	2.159	2.307	2.317	2.209	E6	4.57
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1	1.499	1.520	1.512	1.556	1.688	1.555	E5	4.97
37)	L8	AR-1262-2	2.676	2.833	2.818	2.750	2.842	2.784	E5	2.52
38)	L8	AR-1262-3	1.840	1.839	1.852	1.862	1.944	1.867	E5	2.35
39)	L8	AR-1262-4	1.378	1.372	1.384	1.402	1.459	1.399	E5	2.52
40)	L8	AR-1262-5	1.012	0.993	1.009	1.034	1.054	1.020	E5	2.33
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR035224.D 1600 =PR035226.D 800 =PR035225.D
 200 =PR035223.D 100 =PR035222.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.955	2.304	2.223	1.978	2.187	2.130 E6	7.28	
2) SA	Decachlorobiphenyl	5.170	5.096	5.195	5.308	5.453	5.244 E6	2.65	
3) L1	AR-1016-1						0.000	-1.00	
4) L1	AR-1016-2						0.000	-1.00	
5) L1	AR-1016-3						0.000	-1.00	

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR010419CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jan 05 01:14:53 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035224.D 1600 =PR035226.D 800 =PR035225.D
 200 =PR035223.D 100 =PR035222.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	3.435	3.554	3.516	3.479	3.701	3.537 E5	2.87
37) L8	AR-1262-2	6.789	7.049	7.021	6.783	6.961	6.921 E5	1.83
38) L8	AR-1262-3	2.927	2.822	2.829	2.845	2.962	2.877 E5	2.21
39) L8	AR-1262-4	4.964	5.119	5.104	5.013	5.140	5.068 E5	1.50
40) L8	AR-1262-5	2.471	2.475	2.506	2.522	2.560	2.507 E5	1.45
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range