

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR010419CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jan 05 01:31:33 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035229.D 1600 =PR035231.D 800 =PR035230.D
 200 =PR035228.D 100 =PR035227.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.191	1.101	1.213	0.863	1.049	1.083	E6	12.92
2) SA	Decachlorobiphenyl	4.856	4.686	4.930	4.720	4.897	4.818	E6	2.25
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	5.798	5.539	5.952	5.330	5.635	5.651	E5	4.22
42) L9	AR-1268-2	4.936	4.784	5.088	4.539	4.786	4.827	E5	4.23
43) L9	AR-1268-3	4.475	4.358	4.627	4.149	4.335	4.389	E5	4.04
44) L9	AR-1268-4	1.775	1.712	1.814	1.698	1.762	1.752	E5	2.70
45) L9	AR-1268-5	1.384	1.329	1.425	1.303	1.324	1.353	E6	3.70

Signal #2 Calibration Files

400 =PR035229.D 1600 =PR035231.D 800 =PR035230.D
 200 =PR035228.D 100 =PR035227.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.518	2.291	2.532	1.877	2.391	2.322	E6	11.51
2) SA	Decachlorobiphenyl	1.173	1.142	1.204	1.122	1.154	1.159	E7	2.69
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR010419CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jan 05 01:31:33 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035229.D 1600 =PR035231.D 800 =PR035230.D
 200 =PR035228.D 100 =PR035227.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	1.416	1.338	1.452	1.293	1.354	1.371 E6	4.62
42) L9	AR-1268-2	1.262	1.211	1.303	1.149	1.200	1.225 E6	4.83
43) L9	AR-1268-3	1.105	1.066	1.132	1.003	1.033	1.068 E6	4.89
44) L9	AR-1268-4	4.350	4.263	4.496	4.100	4.268	4.295 E5	3.36
45) L9	AR-1268-5	3.472	3.228	3.531	3.257	3.285	3.355 E6	4.09

(#) = Out of Range