

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010419\
 Data File : PR035216.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jan 2019 21:08
 Operator : SM\SJ
 Sample : AR1248ICC1600
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1248501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 04 23:44:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR010419CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 04 23:40:30 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.430	3.507	83475708	171.5E6	80.942	80.429
2)	SA Decachlor...	10.108	8.393	332.6E6	809.4E6	156.234	158.533
Target Compounds							
21)	L5 AR-1248-1	5.592	4.564	47052261	93826423	1539.667	1541.656
22)	L5 AR-1248-2	5.863	4.794	68776073	133.9E6	1530.172	1532.028
23)	L5 AR-1248-3	6.065	4.834	81200634	139.6E6	1543.716	1537.583
24)	L5 AR-1248-4	6.467	5.002	100.4E6	180.7E6	1533.112	1549.807
25)	L5 AR-1248-5	6.506	5.386	94613866	194.2E6	1545.860	1554.707

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010419\
Data File : PR035216.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jan 2019 21:08
Operator : SM\SJ
Sample : AR1248ICC1600
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1248501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 04 23:44:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR010419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 04 23:40:30 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

