

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010521\
 Data File : PR049211.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jan 2021 19:19
 Operator : DD\AJ
 Sample : AR1248ICC1600
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 06 04:13:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR010521CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 06 04:10:58 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.668	3.853	218.4E6	315.6E6	81.196	78.249
2)	SA Decachlor...	10.555	8.900	361.1E6	1412.0E6	137.914	156.337
Target Compounds							
21)	L5 AR-1248-1	5.843	4.935	90037936	75667201	1428.147	1431.758
22)	L5 AR-1248-2	6.115	5.171	123.9E6	85120706	1364.276	1376.483
23)	L5 AR-1248-3	6.322	5.214	137.3E6	102.8E6	1411.955	1404.597
24)	L5 AR-1248-4	6.726	5.386	167.3E6	136.5E6	1424.051	1433.461
25)	L5 AR-1248-5	6.766	5.779	160.6E6	192.2E6	1430.954	1425.404



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010521\
Data File : PR049211.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jan 2021 19:19
Operator : DD\AJ
Sample : AR1248ICC1600
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 04:13:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR010521CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 06 04:10:58 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#21 AR-1248-1


R.T.: 5.843 min
Delta R.T.: 0.000 min
Response: 90037936
Conc: 1428.15 ng/ml

Instrument :
ECD_R
ClientSampleId :




#21 AR-1248-1

R.T.: 4.935 min
Delta R.T.: 0.000 min
Response: 75667201
Conc: 1431.76 ng/ml




#22 AR-1248-2

R.T.: 6.115 min
Delta R.T.: -0.001 min
Response: 123872353
Conc: 1364.28 ng/ml



#22 AR-1248-2


R.T.: 5.171 min
Delta R.T.: 0.000 min
Response: 85120706
Conc: 1376.48 ng/ml



#23 AR-1248-3


R.T.: 6.322 min
Delta R.T.: -0.001 min
Response: 137324812
Conc: 1411.96 ng/ml

Instrument :
ECD_R
ClientSampleId :




#23 AR-1248-3

R.T.: 5.214 min
Delta R.T.: 0.000 min
Response: 102808124
Conc: 1404.60 ng/ml




#24 AR-1248-4

R.T.: 6.726 min
Delta R.T.: -0.001 min
Response: 167310305
Conc: 1424.05 ng/ml



#24 AR-1248-4


R.T.: 5.386 min
Delta R.T.: 0.000 min
Response: 136478719
Conc: 1433.46 ng/ml



#25 AR-1248-5

R.T.: 6.766 min
Delta R.T.: -0.001 min
Response: 160554867
Conc: 1430.95 ng/ml

Instrument :
ECD_R
ClientSampleId :



#25 AR-1248-5

R.T.: 5.779 min
Delta R.T.: 0.000 min
Response: 192205916
Conc: 1425.40 ng/ml