

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010521\
 Data File : PR049208.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jan 2021 18:33
 Operator : DD\AJ
 Sample : AR1248ICC200
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1248201

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 06 04:11:51 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR010521CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jan 06 04:10:58 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.667	3.852	28712897	42892193	10.674	10.633
2) SA Decachlor...	10.556	8.900	58470118	188.7E6	22.332	20.893
Target Compounds						
21) L5 AR-1248-1	5.844	4.935	14448274	11844310	229.173	224.115
22) L5 AR-1248-2	6.116	5.171	20088812	13344708	221.249	215.797
23) L5 AR-1248-3	6.323	5.214	21527409	15975125	221.342	218.257
24) L5 AR-1248-4	6.727	5.385	26529921	20535216	225.808	215.685
25) L5 AR-1248-5	6.767	5.778	25292183	29229120	225.418	216.764

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010521\
Data File : PR049208.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jan 2021 18:33
Operator : DD\AJ
Sample : AR1248ICC200
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1248201

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 04:11:51 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR010521CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jan 06 04:10:58 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

