

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010523\
 Data File : PR058764.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jan 2023 15:46
 Operator : YP\AJ
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12483342

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 06 03:35:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121422CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 15 03:47:13 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.691	2.913	57594647	67238472	19.095	20.231
2) SA Decachlor...	9.656	8.153	72656164	105.8E6	39.664	38.780
Target Compounds						
21) L5 AR-1248-1	4.924	4.025	19963668	24919622	342.795	335.544
22) L5 AR-1248-2	5.219	4.271	26520880	36534146	364.124	365.977
23) L5 AR-1248-3	5.433	4.313	30467946	36882824	361.720	362.390
24) L5 AR-1248-4	5.869	4.488	31137891	45005515	342.636	359.146
25) L5 AR-1248-5	5.910	4.898	29936083	41991770	339.251	333.493

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010523\
Data File : PR058764.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jan 2023 15:46
Operator : YP\AJ
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12483342

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 03:35:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121422CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 15 03:47:13 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

