

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010523\
 Data File : PR058805.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jan 2023 02:02
 Operator : YP\AJ
 Sample : 01035-11
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 A44D0

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/06/2023
 Supervised By :Ankita Jodhani 01/06/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 06 04:12:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121422CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 15 03:47:13 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.690	2.912	55142934	59088521	18.282	17.779
2)	SA Decachlor...	9.653	8.150	45644443	67326332	24.918	24.673
Target Compounds							
16)	L4 AR-1242-1	4.928	4.028	22540351	33845688	299.472m	347.440m
17)	L4 AR-1242-2	4.946	4.042	221.4E6	329.3E6	2164.307m	2461.963m
18)	L4 AR-1242-3	5.023	4.221	62264383	235.5E6	932.124	3279.120 #
19)	L4 AR-1242-4	5.113	4.311	144.2E6	558.6E6	2674.731	8204.312 #
20)	L4 AR-1242-5	5.908	4.855	280.9E6	240.7E6	5332.983	2694.969 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010523\
Data File : PR058805.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jan 2023 02:02
Operator : YP\AJ
Sample : 01035-11
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
A44D0

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/06/2023
Supervised By :Ankita Jodhani 01/06/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 04:12:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121422CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 15 03:47:13 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

