


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010623\
Data File : PR058828.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jan 2023 12:53
Operator : YP\AJ
Sample : 01034-15DL2 100X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 21:17:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121422CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 15 03:47:13 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) AR-1242-1 (L4)

R.T.	Response	Conc
4.93	2297038	30.52
4.95	18402672	179.88
5.02	8668490	129.77
5.11	36239036	672.29
5.91	38595542	732.63


(16) AR-1242-1 #2 (L4)

R.T.	Response	Conc
4.04	15357860	157.65
4.04	15357860	114.83
4.22	50228771	699.29
4.31	123445740	1812.94
4.86	37879846	424.17

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010623\
Data File : PR058828.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jan 2023 12:53
Operator : YP\AJ
Sample : 01034-15DL2 100X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 06 21:17:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121422CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 15 03:47:13 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) AR-1242-1 #2 (L4)
R.T. Response Conc
4.93 2297038 30.52
4.95 18402672 179.88
5.02 8668490 129.77
5.11 36239036 672.29
5.91 38595542 732.63

(16) AR-1242-1 #2 (L4)
R.T. Response Conc
4.03 2519483 25.86
4.04 12617957 94.34
4.22 50228771 699.29
4.31 123445740 1812.94
4.86 37879846 424.17