

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010924\
 Data File : PR064878.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jan 2024 14:30
 Operator : AJ\MA
 Sample : P1041-14
 Misc :
 ALS Vial : 7 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 PCB-GCP2-BLANK-SPIKE

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 09 21:48:25 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121423CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 15 04:19:57 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 01/10/2024
 Supervised By :Ankita Jodhani 01/10/2024

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 (L7)
 R.T. Response Conc
 6.69 99804616 399.39
 6.98 115803017 405.12
 7.38 72777066 343.24
 7.59 13573614 59.05
 7.97 165204718 385.02

(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 5.75 118648757 410.58
 5.95 143155519 412.04
 6.12 146319897 444.54
 6.63 99316065 368.22
 6.90 245076038 392.17

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR010924\
Data File : PR064878.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jan 2024 14:30
Operator : AJ\MA
Sample : P1041-14
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
PCB-GCP2-BLANK-SPIKE

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 01/10/2024
Supervised By :Ankita Jodhani 01/10/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 09 21:48:25 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121423CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 15 04:19:57 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 (L7)
R.T. Response Conc
6.69 99804616 399.39
6.98 115803017 405.12
7.38 72777066 343.24
7.62 85919684 373.80
7.97 165204718 385.02

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.75 118648757 410.58
5.95 143155519 412.04
6.12 146319897 444.54
6.63 99316065 368.22
6.90 245076038 392.17

(+) = Expected Retention Time