

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011224\
 Data File : PR064909.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jan 2024 22:54
 Operator : AJ\MA
 Sample : AR1254ICC200
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12542202

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 13 03:57:31 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 13 03:55:27 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.703	3.006	67420588	61887744	10.798	10.591
2) SA Decachlor...	9.745	8.384	54449183	62448408	22.110	21.561
Target Compounds						
26) L6 AR-1254-1	5.866	5.009	47815999	65782314	222.640	215.546
27) L6 AR-1254-2	6.107	5.172	71274577	57084608	220.110	216.774
28) L6 AR-1254-3	6.505	5.602	71833867	91310221	217.985	213.321
29) L6 AR-1254-4	6.821	5.853	49997693	56940336	221.955	213.090
30) L6 AR-1254-5	7.282	6.306	56394545	81270531	219.074	213.545



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011224\
Data File : PR064909.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jan 2024 22:54
Operator : AJ\MA
Sample : AR1254ICC200
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12542202

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 13 03:57:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 13 03:55:27 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#26 AR-1254-1


R.T.: 5.866 min
Delta R.T.: 0.000 min
Response: 47815999
Conc: 222.64 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR12542202




#26 AR-1254-1

R.T.: 5.009 min
Delta R.T.: 0.000 min
Response: 65782314
Conc: 215.55 ng/ml




#27 AR-1254-2

R.T.: 6.107 min
Delta R.T.: 0.000 min
Response: 71274577
Conc: 220.11 ng/ml



#27 AR-1254-2


R.T.: 5.172 min
Delta R.T.: 0.000 min
Response: 57084608
Conc: 216.77 ng/ml



#28 AR-1254-3


R.T.: 6.505 min
Delta R.T.: 0.001 min
Response: 71833867
Conc: 217.99 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR12542202




#28 AR-1254-3

R.T.: 5.602 min
Delta R.T.: 0.000 min
Response: 91310221
Conc: 213.32 ng/ml




#29 AR-1254-4

R.T.: 6.821 min
Delta R.T.: 0.001 min
Response: 49997693
Conc: 221.96 ng/ml



#29 AR-1254-4


R.T.: 5.853 min
Delta R.T.: 0.000 min
Response: 56940336
Conc: 213.09 ng/ml



#30 AR-1254-5

R.T.: 7.282 min
Delta R.T.: 0.001 min
Response: 56394545
Conc: 219.07 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR12542202



#30 AR-1254-5

R.T.: 6.306 min
Delta R.T.: 0.000 min
Response: 81270531
Conc: 213.55 ng/ml