

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011224\
 Data File : PR064901.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jan 2024 20:55
 Operator : AJ\MA
 Sample : AR1242ICC800
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12424244

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 13 03:13:23 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 13 03:13:15 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.703	3.006	250.7E6	231.8E6	38.240	39.635
2)	SA Decachlor...	9.743	8.382	182.8E6	218.4E6	72.433	75.415
Target Compounds							
16)	L4 AR-1242-1	4.944	4.152	115.0E6	120.7E6	730.713	757.445
17)	L4 AR-1242-2	4.967	4.171	170.4E6	166.4E6	739.656	769.341
18)	L4 AR-1242-3	5.032	4.355	106.9E6	90371348	723.058	763.737
19)	L4 AR-1242-4	5.136	4.449	86582320	86159319	737.773	747.571
20)	L4 AR-1242-5	5.936	5.008	91541120	110.4E6	725.992	755.231

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011224\
Data File : PR064901.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jan 2024 20:55
Operator : AJ\MA
Sample : AR1242ICC800
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12424244

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 13 03:13:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 13 03:13:15 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

