

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011224\
 Data File : PR064900.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jan 2024 20:40
 Operator : AJ\MA
 Sample : AR1242ICC400
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12423243

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 13 03:01:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 13 02:59:56 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.703	3.005	125.9E6	114.7E6	20.000	20.000
2) SA Decachlor...	9.743	8.382	95797439	111.4E6	40.000	40.000
Target Compounds						
16) L4 AR-1242-1	4.944	4.152	59482078	61456870	400.000	400.000
17) L4 AR-1242-2	4.966	4.171	88322944	84284633	400.000	400.000
18) L4 AR-1242-3	5.031	4.355	56325773	46031817	400.000	400.000
19) L4 AR-1242-4	5.135	4.449	45220276	44498103	400.000	400.000
20) L4 AR-1242-5	5.936	5.009	47723658	56476427	400.000	400.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011224\
Data File : PR064900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jan 2024 20:40
Operator : AJ\MA
Sample : AR1242ICC400
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12423243

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 13 03:01:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 13 02:59:56 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

