

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011323\
 Data File : PR058936.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2023 05:16
 Operator : YP\AJ
 Sample : 01145-18
 Misc :
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 A44B7

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/16/2023
 Supervised By :Ankita Jodhani 01/16/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 06:44:28 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 14 03:30:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.691	2.912	73408433	86505537	19.663	21.589
2)	SA Decachlor...	9.659	8.150	84976929	121.2E6	38.305	36.476
Target Compounds							
16)	L4 AR-1242-1	4.930	4.028	1470427	1085494	18.357m	10.933m#
17)	L4 AR-1242-2	4.946	4.042	6465667	5957514	57.402m	41.314m#
18)	L4 AR-1242-3	5.023	4.220	1586442	4569367	21.689m	60.413 #
19)	L4 AR-1242-4	5.111	4.310	4267014	11284832	73.357m	151.007 #
20)	L4 AR-1242-5	5.909	4.854	6476183	6644643	120.183	72.223 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011323\
Data File : PR058936.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2023 05:16
Operator : YP\AJ
Sample : 01145-18
Misc :
ALS Vial : 69 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
A44B7

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/16/2023
Supervised By :Ankita Jodhani 01/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 14 06:44:28 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 14 03:30:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

