

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011323\
 Data File : PR058919.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2023 01:06
 Operator : YP\AJ
 Sample : 01145-08
 Misc :
 ALS Vial : 52 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 A43W5

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 03:51:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 14 03:30:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 01/16/2023
 Supervised By :Ankita Jodhani 01/16/2023

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) AR-1242-1 (L4)
 R.T. Response Conc
 4.96 19019547376 237436.38
 4.96 19019547376 168855.01
 5.02 8810538174 120450.71
 5.11 25256708144 434202.17
 5.91 28485515772 528627.60

(16) AR-1242-1 #2 (L4)
 R.T. Response Conc
 4.04 8760705360 88240.39
 4.04 8760705360 60752.86
 4.23 34839158554 460618.83
 4.32 47988392564 642152.68
 4.86 30825184314 335047.68

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011323\
 Data File : PR058919.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2023 01:06
 Operator : YP\AJ
 Sample : 01145-08
 Misc :
 ALS Vial : 52 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 A43W5

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 01/16/2023
 Supervised By :Ankita Jodhani 01/16/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 14 03:51:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 14 03:30:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) AR-1242-1 #2 (L4)
 R.T. Response Conc
 4.93 747437826 9330.87
 4.96 17891400494 158839.35
 5.02 8810538174 120450.71
 5.11 25256708144 434202.17
 5.91 28485515772 528627.60

(16) AR-1242-1 #2 (L4)
 R.T. Response Conc
 4.03 1296608029 13059.82
 4.04 7180226500 49792.72
 4.23 34839158554 460618.83
 4.32 47988392564 642152.68
 4.86 30825184314 335047.68