

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011423\
 Data File : PR058966.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2023 13:00
 Operator : YP\AJ
 Sample : 01145-08DL 100X
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 A43W5DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/16/2023
 Supervised By :Ankita Jodhani 01/16/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 16 04:23:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 14 03:30:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds						
16)	L4	AR-1242-1	4.924	4.024	15608797 20535389	194.857 206.838m
17)	L4	AR-1242-2	4.961	4.041	403.8E6 112.9E6 3584.918	783.057m#
18)	L4	AR-1242-3	5.023	4.221	184.2E6 1117.3E6 2518.469	14772.386 #
19)	L4	AR-1242-4	5.111	4.310	609.3E6 2978.7E6 10474.423	39858.566 #
20)	L4	AR-1242-5	5.909	4.854	659.8E6 747.3E6 12244.617	8122.232 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011423\
Data File : PR058966.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2023 13:00
Operator : YP\AJ
Sample : 01145-08DL 100X
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
A43W5DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/16/2023
Supervised By :Ankita Jodhani 01/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 16 04:23:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 14 03:30:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

