

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011423\
 Data File : PR058968.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jan 2023 13:29
 Operator : YP\AJ
 Sample : 01145-21DL 10X
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 A44C9DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/16/2023
 Supervised By :Ankita Jodhani 01/16/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 16 04:24:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011323CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 14 03:30:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.689	2.912	7014070	7451852	1.879	1.860
2) SA Decachlor...	9.658	8.150	9089033	11581201	4.097	3.487
Target Compounds						
16) L4 AR-1242-1	4.924	4.024	3393694	3603099	42.366	36.291m
17) L4 AR-1242-2	4.954	4.042	24759699	17050620	219.816	118.241m#
18) L4 AR-1242-3	5.022	4.221	12252829	74488691	167.511	984.837 #
19) L4 AR-1242-4	5.111	4.310	49656356	200.3E6	853.670	2679.749 #
20) L4 AR-1242-5	5.909	4.854	69426794	65036822	1288.406	706.904 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011423\
Data File : PR058968.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jan 2023 13:29
Operator : YP\AJ
Sample : 01145-21DL 10X
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
A44C9DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/16/2023
Supervised By :Ankita Jodhani 01/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 16 04:24:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011323CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 14 03:30:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

