

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR011723\  
Data File : PR059053.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 17 Jan 2023 19:45  
Operator : YP\AJ  
Sample : AIBLK83  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
ECD\_R  
ClientSampleId :  
AIBLK528

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jan 17 21:01:31 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR011323CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Sat Jan 14 03:30:33 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound        | RT#1  | RT#2  | Resp#1   | Resp#2   | ng/ml  | ng/ml  |
|-----------------------------|-----------------|-------|-------|----------|----------|--------|--------|
| -----                       |                 |       |       |          |          |        |        |
| System Monitoring Compounds |                 |       |       |          |          |        |        |
| 1)                          | SA Tetrachlo... | 3.689 | 2.911 | 67306150 | 77773089 | 18.028 | 19.410 |
| 2)                          | SA Decachlor... | 9.658 | 8.152 | 87627981 | 129.3E6  | 39.500 | 38.923 |

Target Compounds

-----

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR011723\  
Data File : PR059053.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 17 Jan 2023 19:45  
Operator : YP\AJ  
Sample : AIBLK83  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
ECD\_R  
ClientSampleId :  
AIBLK528

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jan 17 21:01:31 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR011323CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Sat Jan 14 03:30:33 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

