

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011924\
 Data File : PR065009.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jan 2024 17:03
 Operator : AJ\MA
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 19 20:22:02 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 19 20:16:29 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.702	3.005	623.4E6	613.2E6	100.183	101.068
2)	SA Decachlor...	9.745	8.384	688.4E6	876.9E6	99.032	99.775
Target Compounds							
41)	L9 AR-1268-1	8.273	7.208	588.6E6	786.3E6	996.475	993.500
42)	L9 AR-1268-2	8.367	7.278	533.0E6	750.6E6	995.883	1001.834
43)	L9 AR-1268-3	8.593	7.504	461.3E6	636.2E6	996.952	1001.605
44)	L9 AR-1268-4	9.016	7.823	193.2E6	259.9E6	984.854	1000.688
45)	L9 AR-1268-5	9.420	8.123	1454.3E6	1952.3E6	1008.625	1005.984



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR011924\
Data File : PR065009.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jan 2024 17:03
Operator : AJ\MA
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 19 20:22:02 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011924.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 19 20:16:29 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.702 min
Delta R.T.: 0.000 min
Response: 623373952
Conc: 100.18 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1268ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.005 min
Delta R.T.: 0.000 min
Response: 613223896
Conc: 101.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.745 min
Delta R.T.: 0.000 min
Response: 688378491
Conc: 99.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.384 min
Delta R.T.: 0.000 min
Response: 876889766
Conc: 99.77 ng/ml



#41 AR-1268-1


R.T.: 8.273 min
Delta R.T.: 0.000 min
Response: 588560328
Conc: 996.48 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1268ICC1000




#41 AR-1268-1

R.T.: 7.208 min
Delta R.T.: 0.000 min
Response: 786263065
Conc: 993.50 ng/ml




#42 AR-1268-2

R.T.: 8.367 min
Delta R.T.: 0.000 min
Response: 533010292
Conc: 995.88 ng/ml



#42 AR-1268-2


R.T.: 7.278 min
Delta R.T.: 0.000 min
Response: 750570447
Conc: 1001.83 ng/ml



#43 AR-1268-3


R.T.: 8.593 min
Delta R.T.: 0.000 min
Response: 461339585
Conc: 996.95 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1268ICC1000




#43 AR-1268-3

R.T.: 7.504 min
Delta R.T.: 0.000 min
Response: 636168359
Conc: 1001.61 ng/ml




#44 AR-1268-4

R.T.: 9.016 min
Delta R.T.: 0.000 min
Response: 193189515
Conc: 984.85 ng/ml



#44 AR-1268-4


R.T.: 7.823 min
Delta R.T.: 0.000 min
Response: 259937871
Conc: 1000.69 ng/ml



#45 AR-1268-5

R.T.: 9.420 min
Delta R.T.: 0.000 min
Response: 1454250208
Conc: 1008.63 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1268ICC1000



#45 AR-1268-5

R.T.: 8.123 min
Delta R.T.: 0.000 min
Response: 1952250312
Conc: 1005.98 ng/ml