

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012024\
 Data File : PR065036.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jan 2024 23:43
 Operator : AJ\MA
 Sample : P1157-10
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AC0

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/23/2024
 Supervised By :Sohil Jodhani 01/23/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 20 02:17:39 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 19 21:16:50 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.690	3.005	1069.4E6	81491389	164.498	14.035m#
2)	SA Decachlor...	9.745	8.382	161.5E6	156.6E6	66.522	55.758
Target Compounds							
16)	L4 AR-1242-1	4.944	4.152	5052.6E6	5240.1E6	33105.990	33652.371
17)	L4 AR-1242-2	4.979	4.171	12813.9E6	6262.1E6	57184.807	29522.539 #
18)	L4 AR-1242-3	5.045	4.353	7081.0E6	1776.8E6	49646.347	15353.526 #
19)	L4 AR-1242-4	5.136	4.448	1423.5E6	19052.8E6	12508.212	170105.328 #
20)	L4 AR-1242-5	5.939	5.009	8111.5E6	10017.2E6	66601.055	70142.541
26)	L6 AR-1254-1	5.867	5.009	7878.2E6	10017.2E6	36013.345	32385.733
27)	L6 AR-1254-2	6.107	5.172	13231.6E6	10447.6E6	40221.770	39054.766
28)	L6 AR-1254-3	6.505	5.602	14159.3E6	18312.9E6	42342.000	42487.251
29)	L6 AR-1254-4	6.821	5.852	8218.4E6	8359.5E6	35914.281	31100.148
30)	L6 AR-1254-5	7.281	6.304	10297.6E6	12501.6E6	39536.659	32527.128

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012024\
Data File : PR065036.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jan 2024 23:43
Operator : AJ\MA
Sample : P1157-10
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0AC0

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/23/2024
Supervised By :Sohil Jodhani 01/23/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 20 02:17:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 19 21:16:50 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

