

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012120\
 Data File : PR044459.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jan 2020 13:51
 Operator : DD\AJ
 Sample : L1161-09
 Misc : AR1248 MDL 25 PPB
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-WATER-03-QT1-2020

Manual Integrations
 APPROVED

Ankita
 1/22/2020 10:59:02 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 01:19:46 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012020.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jan 21 02:46:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.699	3.962	162.0E6	330.5E6	21.845	20.899
2) SA Decachlor...	10.592	9.039	164.0E6	286.9E6	21.463	20.159
Target Compounds						
21) L5 AR-1248-1	5.874	5.051	4941370	7038977	32.582	27.901m
22) L5 AR-1248-2	6.145	5.288	6052175	10436843	29.688m	38.095m#
23) L5 AR-1248-3	6.353	5.330	6638022	10274814	29.153m	36.639m#
24) L5 AR-1248-4	6.756	5.504	9173804	11732133	33.887m	32.468
25) L5 AR-1248-5	6.795	5.896	8394920	12140608	32.806m	28.952m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012120\
Data File : PR044459.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jan 2020 13:51
Operator : DD\AJ
Sample : L1161-09
Misc : AR1248 MDL 25 PPB
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-03-QT1-2020

Manual Integrations
APPROVED

Ankita
1/22/2020 10:59:02 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 01:19:46 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012020.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jan 21 02:46:28 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

