

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012224\
 Data File : PR065085.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jan 2024 14:34
 Operator : AJ\MA
 Sample : P1157-02DL2 20X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AA9DL2

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/23/2024
 Supervised By :Sohil Jodhani 01/23/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 22 21:48:41 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 19 21:16:50 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.702	3.005	2979653	2427732	0.456	0.417
2)	SA Decachlor...	9.744	8.375	3022423	2602473	1.242m	0.922m#
Target Compounds							
16)	L4 AR-1242-1	4.944	4.152	35235699	38439322	230.873	246.860
17)	L4 AR-1242-2	4.966	4.170	95395182	95764156	425.720	451.479
18)	L4 AR-1242-3	5.033	4.353	20837361	47256791	146.094	408.345 #
19)	L4 AR-1242-4	5.135	4.447	43155121	47281546	379.215	422.133
20)	L4 AR-1242-5	5.936	5.007	52875070	114.4E6	434.143	801.261 #
26)	L6 AR-1254-1	5.866	5.007	70868621	114.4E6	323.958	369.953
27)	L6 AR-1254-2	6.106	5.169	113.8E6	86079066	345.918	321.777
28)	L6 AR-1254-3	6.504	5.600	105.2E6	144.6E6	314.572	335.464
29)	L6 AR-1254-4	6.819	5.850	70516562	84050831	308.156m	312.696
30)	L6 AR-1254-5	7.280	6.303	90450909	138.3E6	347.279	359.911

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012224\
Data File : PR065085.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jan 2024 14:34
Operator : AJ\MA
Sample : P1157-02DL2 20X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0AA9DL2

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/23/2024
Supervised By :Sohil Jodhani 01/23/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 22 21:48:41 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 19 21:16:50 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

