

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012224\
 Data File : PR065131.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jan 2024 02:10
 Operator : AJ\MA
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12543331

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/23/2024
 Supervised By :Sohil Jodhani 01/23/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 03:57:26 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 19 21:16:50 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.702	3.005	116.8E6	110.3E6	17.874	18.943
2) SA Decachlor...	9.737	8.374	89224202	99069162	36.671	35.101
Target Compounds						
26) L6 AR-1254-1	5.865	5.006	76899356	117.3E6	351.526	379.260
27) L6 AR-1254-2	6.106	5.168	117.7E6	101.9E6	357.852	380.934
28) L6 AR-1254-3	6.503	5.599	121.5E6	162.6E6	363.233	377.325
29) L6 AR-1254-4	6.819	5.849	83367303	100.0E6	364.314m	372.107
30) L6 AR-1254-5	7.279	6.302	94341088	140.1E6	362.215	364.438

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012224\
Data File : PR065131.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jan 2024 02:10
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12543331

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/23/2024
Supervised By :Sohil Jodhani 01/23/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 03:57:26 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 19 21:16:50 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

