

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012224\
 Data File : PR065134.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jan 2024 02:54
 Operator : AJ\MA
 Sample : P1158-11 4X
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AK7

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/23/2024
 Supervised By :Sohil Jodhani 01/23/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 23 03:59:08 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 19 21:16:50 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.701	3.005	31359575	18923732	4.800	3.250 #
2)	SA Decachlor...	9.739	8.375	96480805	91541150	39.653	32.434
Target Compounds							
16)	L4 AR-1242-1	4.944	4.151	202.8E6	196.2E6	1328.830	1259.921
17)	L4 AR-1242-2	4.967	4.169	641.2E6	517.2E6	2861.655	2438.122
18)	L4 AR-1242-3	5.039	4.353	104.9E6	211.6E6	735.208	1828.691 #
19)	L4 AR-1242-4	5.135	4.447	186.5E6	651.1E6	1638.843	5812.791 #
20)	L4 AR-1242-5	5.936	5.006	874.4E6	2175.8E6	7179.552	15235.589 #
26)	L6 AR-1254-1	5.866	5.006	1551.6E6	2175.8E6	7092.529	7034.471
27)	L6 AR-1254-2	6.107	5.169	2687.7E6	1886.2E6	8170.278	7050.883
28)	L6 AR-1254-3	6.503	5.599	3304.1E6	3219.4E6	9880.595	7469.231
29)	L6 AR-1254-4	6.819	5.850	1921.3E6	1925.4E6	8396.096m	7163.093
30)	L6 AR-1254-5	7.281	6.303	2419.2E6	6025.7E6	9288.355	15677.962 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012224\
Data File : PR065134.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jan 2024 02:54
Operator : AJ\MA
Sample : P1158-11 4X
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0AK7

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 01/23/2024
Supervised By :Sohil Jodhani 01/23/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 23 03:59:08 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 19 21:16:50 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

