

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012620\
 Data File : PR044544.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jan 2020 16:36
 Operator : AJ\MA
 Sample : AR1248ICC800
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1248401

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 24 17:13:32 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012620CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 24 17:12:14 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.693	3.955	289.0E6	590.1E6	41.782	40.377
2) SA Decachlor...	10.583	9.032	598.3E6	1107.0E6	77.002	78.241
Target Compounds						
21) L5 AR-1248-1	5.868	5.046	125.2E6	179.4E6	783.491	753.946
22) L5 AR-1248-2	6.141	5.283	167.8E6	220.0E6	777.096	756.541
23) L5 AR-1248-3	6.348	5.326	190.7E6	229.8E6	786.055	759.075
24) L5 AR-1248-4	6.751	5.499	230.7E6	309.5E6	784.148	774.586
25) L5 AR-1248-5	6.790	5.892	218.7E6	365.1E6	783.496	770.870

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012620\
Data File : PR044544.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jan 2020 16:36
Operator : AJ\MA
Sample : AR1248ICC800
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleID :
AR1248401

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 24 17:13:32 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012620CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 24 17:12:14 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

