

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR012624\  
 Data File : PR065214.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 25 Jan 2024 19:11  
 Operator : AJ\MA  
 Sample : P1187-07  
 Misc : AR1262 LOD 25 PPB  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 ECD\_R  
 ClientSampleId :  
 LOD-MDL-WATER-01-QT1-2024

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jan 26 02:25:01 2024  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR011924.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Fri Jan 19 20:49:45 2024  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound        | RT#1   | RT#2   | Resp#1   | Resp#2   | ng/ml  | ng/ml    |
|-----------------------------|-----------------|--------|--------|----------|----------|--------|----------|
| -----                       |                 |        |        |          |          |        |          |
| System Monitoring Compounds |                 |        |        |          |          |        |          |
| 1)                          | SA Tetrachlo... | 3.701  | 3.003  | 118.4E6  | 109.0E6  | 19.843 | 19.922   |
| 2)                          | SA Decachlor... | 9.738  | 8.374  | 89358535 | 95895355 | 22.342 | 20.262   |
| Target Compounds            |                 |        |        |          |          |        |          |
| 3)                          | L1 AR-1016-1    | 4.942  | 4.148  | 257066   | 414753   | 1.450  | 2.284 #  |
| 4)                          | L1 AR-1016-2    | 4.965  | 4.167  | 317250   | 217313   | 1.217  | 0.879 #  |
| 5)                          | L1 AR-1016-3    | 5.027  | 4.352  | 250485   | 188721   | 1.499  | 1.392    |
| 6)                          | L1 AR-1016-4    | 5.139  | 4.405  | 385307   | 336733   | 2.874  | 3.215    |
| 7)                          | L1 AR-1016-5    | 0.000  | 4.622  | 0        | 426847   | N.D.   | 3.060 #  |
| 8)                          | L2 AR-1221-1    | 3.912f | 3.228  | 3057016  | 108788   | 42.593 | 1.757 #  |
| 9)                          | L2 AR-1221-2    | 4.016  | 3.313  | 1672778  | 1695959  | 31.909 | 35.558   |
| 10)                         | L2 AR-1221-3    | 4.068f | 3.400  | 843460   | 400597   | 5.490  | 2.900 #  |
| 11)                         | L3 AR-1232-1    | 4.068f | 3.400  | 843460   | 400597   | 6.240  | 3.345 #  |
| 12)                         | L3 AR-1232-2    | 4.636f | 4.167  | 804023   | 217313   | 10.436 | 1.779 #  |
| 13)                         | L3 AR-1232-3    | 4.965  | 4.352  | 317250   | 188721   | 2.488  | 2.837    |
| 14)                         | L3 AR-1232-4    | 5.139  | 4.444  | 385307   | 244586   | 5.878  | 4.271 #  |
| 15)                         | L3 AR-1232-5    | 5.240  | 4.622  | 588328   | 426847   | 12.164 | 6.696 #  |
| 16)                         | L4 AR-1242-1    | 4.942  | 4.148  | 257066   | 414753   | 1.798  | 2.838 #  |
| 17)                         | L4 AR-1242-2    | 4.965  | 4.167  | 317250   | 217313   | 1.506  | 1.089 #  |
| 18)                         | L4 AR-1242-3    | 5.027  | 4.352  | 250485   | 188721   | 1.850  | 1.724    |
| 19)                         | L4 AR-1242-4    | 5.139  | 4.444  | 385307   | 244586   | 3.546  | 2.330 #  |
| 20)                         | L4 AR-1242-5    | 5.935  | 5.003  | 453554   | 1008855  | 4.010  | 7.472 #  |
| 21)                         | L5 AR-1248-1    | 4.942  | 4.148  | 257066   | 414753   | 2.416  | 3.840 #  |
| 22)                         | L5 AR-1248-2    | 5.240  | 4.405  | 588328   | 336733   | 3.786  | 2.325 #  |
| 23)                         | L5 AR-1248-3    | 0.000  | 4.444  | 0        | 244586   | N.D.   | 1.581 #  |
| 24)                         | L5 AR-1248-4    | 5.891  | 4.622  | 197150   | 426847   | 1.040  | 2.279 #  |
| 25)                         | L5 AR-1248-5    | 5.935  | 5.029f | 453554   | 656603   | 2.380  | 3.528 #  |
| 26)                         | L6 AR-1254-1    | 5.866  | 5.003  | 1414772  | 1008855  | 6.908  | 3.462 #  |
| 27)                         | L6 AR-1254-2    | 6.107  | 5.165  | 1260835  | 1135326  | 4.045  | 4.482    |
| 28)                         | L6 AR-1254-3    | 6.493  | 5.616f | 1602025  | 3991696  | 5.008  | 9.756 #  |
| 29)                         | L6 AR-1254-4    | 6.812  | 5.848  | 1605509  | 553758   | 7.006  | 2.157 #  |
| 30)                         | L6 AR-1254-5    | 7.277  | 6.298  | 5523446  | 4777342  | 22.224 | 13.033 # |
| 31)                         | L7 AR-1260-1    | 6.685  | 5.739  | 6393769  | 5711993  | 25.435 | 20.011   |
| 32)                         | L7 AR-1260-2    | 6.972  | 5.947  | 6426918  | 6372995  | 22.363 | 18.378   |
| 33)                         | L7 AR-1260-3    | 7.368  | 6.111  | 8106399  | 10543686 | 38.343 | 32.453   |
| 34)                         | L7 AR-1260-4    | 7.612  | 6.623  | 8084991  | 7370674  | 34.262 | 27.899   |
| 35)                         | L7 AR-1260-5    | 7.957  | 6.892  | 12429406 | 14747724 | 28.269 | 23.913   |
| 36)                         | L8 AR-1262-1    | 7.368  | 6.401  | 8106399  | 3456506  | 23.948 | 19.270   |
| 37)                         | L8 AR-1262-2    | 7.957  | 6.623  | 12429406 | 7370674  | 21.974 | 18.943   |
| 38)                         | L8 AR-1262-3    | 8.272  | 7.202  | 8260976  | 6525594  | 21.709 | 23.023   |
| 39)                         | L8 AR-1262-4    | 8.359  | 7.271  | 6151393  | 11561781 | 21.000 | 20.076   |
| 40)                         | L8 AR-1262-5    | 9.009  | 7.816  | 5484376  | 5991192  | 27.368 | 23.019   |
| 41)                         | L9 AR-1268-1    | 8.272  | 7.202  | 8260976  | 6525594  | 13.626 | 8.226 #  |

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR012624\  
Data File : PR065214.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 25 Jan 2024 19:11  
Operator : AJ\MA  
Sample : P1187-07  
Misc : AR1262 LOD 25 PPB  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jan 26 02:25:01 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR011924.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Fri Jan 19 20:49:45 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

|     | Compound     | RT#1  | RT#2  | Resp#1  | Resp#2   | ng/ml  | ng/ml    |
|-----|--------------|-------|-------|---------|----------|--------|----------|
| 42) | L9 AR-1268-2 | 8.359 | 7.271 | 6151393 | 11561781 | 11.242 | 15.530 # |
| 43) | L9 AR-1268-3 | 8.584 | 7.498 | 1724179 | 4045754  | 3.634  | 6.386 #  |
| 44) | L9 AR-1268-4 | 9.009 | 7.816 | 5484376 | 5991192  | 26.953 | 22.840   |
| 45) | L9 AR-1268-5 | 9.412 | 8.115 | 2090862 | 2811182  | 1.434  | 1.468    |



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR012624\  
Data File : PR065214.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 25 Jan 2024 19:11  
Operator : AJ\MA  
Sample : P1187-07  
Misc : AR1262 LOD 25 PPB  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jan 26 02:25:01 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR011924.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Fri Jan 19 20:49:45 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.701 min  
Delta R.T.: -0.002 min  
Response: 118449420  
Conc: 19.84 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#1 Tetrachloro-m-xylene

R.T.: 3.003 min  
Delta R.T.: -0.001 min  
Response: 108971593  
Conc: 19.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.738 min  
Delta R.T.: -0.004 min  
Response: 89358535  
Conc: 22.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.374 min  
Delta R.T.: -0.007 min  
Response: 95895355  
Conc: 20.26 ng/ml



#3 AR-1016-1


R.T.: 4.942 min  
Delta R.T.: -0.003 min  
Response: 257066  
Conc: 1.45 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#3 AR-1016-1

R.T.: 4.148 min  
Delta R.T.: -0.004 min  
Response: 414753  
Conc: 2.28 ng/ml




#4 AR-1016-2

R.T.: 4.965 min  
Delta R.T.: -0.002 min  
Response: 317250  
Conc: 1.22 ng/ml



#4 AR-1016-2


R.T.: 4.167 min  
Delta R.T.: -0.003 min  
Response: 217313  
Conc: 0.88 ng/ml



#5 AR-1016-3


R.T.: 5.027 min  
Delta R.T.: -0.005 min  
Response: 250485  
Conc: 1.50 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#5 AR-1016-3

R.T.: 4.352 min  
Delta R.T.: -0.001 min  
Response: 188721  
Conc: 1.39 ng/ml




#6 AR-1016-4

R.T.: 5.139 min  
Delta R.T.: 0.004 min  
Response: 385307  
Conc: 2.87 ng/ml



#6 AR-1016-4


R.T.: 4.405 min  
Delta R.T.: 0.000 min  
Response: 336733  
Conc: 3.22 ng/ml



#7 AR-1016-5


R.T.: 0.000 min  
Exp R.T.: 5.456 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#7 AR-1016-5

R.T.: 4.622 min  
Delta R.T.: -0.006 min  
Response: 426847  
Conc: 3.06 ng/ml




#8 AR-1221-1

R.T.: 3.912 min  
Delta R.T.: -0.015 min  
Response: 3057016  
Conc: 42.59 ng/ml



#8 AR-1221-1


R.T.: 3.228 min  
Delta R.T.: -0.007 min  
Response: 108788  
Conc: 1.76 ng/ml



#9 AR-1221-2


R.T.: 4.016 min  
Delta R.T.: -0.001 min  
Response: 1672778  
Conc: 31.91 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#9 AR-1221-2

R.T.: 3.313 min  
Delta R.T.: -0.010 min  
Response: 1695959  
Conc: 35.56 ng/ml




#10 AR-1221-3

R.T.: 4.068 min  
Delta R.T.: -0.029 min  
Response: 843460  
Conc: 5.49 ng/ml



#10 AR-1221-3


R.T.: 3.400 min  
Delta R.T.: -0.002 min  
Response: 400597  
Conc: 2.90 ng/ml



#11 AR-1232-1


R.T.: 4.068 min  
Delta R.T.: -0.029 min  
Response: 843460  
Conc: 6.24 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#11 AR-1232-1

R.T.: 3.400 min  
Delta R.T.: -0.003 min  
Response: 400597  
Conc: 3.34 ng/ml




#12 AR-1232-2

R.T.: 4.636 min  
Delta R.T.: -0.018 min  
Response: 804023  
Conc: 10.44 ng/ml



#12 AR-1232-2


R.T.: 4.167 min  
Delta R.T.: -0.003 min  
Response: 217313  
Conc: 1.78 ng/ml



#13 AR-1232-3


R.T.: 4.965 min  
Delta R.T.: -0.002 min  
Response: 317250  
Conc: 2.49 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#13 AR-1232-3

R.T.: 4.352 min  
Delta R.T.: -0.001 min  
Response: 188721  
Conc: 2.84 ng/ml




#14 AR-1232-4

R.T.: 5.139 min  
Delta R.T.: 0.004 min  
Response: 385307  
Conc: 5.88 ng/ml



#14 AR-1232-4


R.T.: 4.444 min  
Delta R.T.: -0.004 min  
Response: 244586  
Conc: 4.27 ng/ml



#15 AR-1232-5


R.T.: 5.240 min  
Delta R.T.: 0.000 min  
Response: 588328  
Conc: 12.16 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#15 AR-1232-5

R.T.: 4.622 min  
Delta R.T.: -0.007 min  
Response: 426847  
Conc: 6.70 ng/ml




#16 AR-1242-1

R.T.: 4.942 min  
Delta R.T.: -0.002 min  
Response: 257066  
Conc: 1.80 ng/ml



#16 AR-1242-1


R.T.: 4.148 min  
Delta R.T.: -0.004 min  
Response: 414753  
Conc: 2.84 ng/ml



#17 AR-1242-2


R.T.: 4.965 min  
Delta R.T.: -0.002 min  
Response: 317250  
Conc: 1.51 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#17 AR-1242-2

R.T.: 4.167 min  
Delta R.T.: -0.003 min  
Response: 217313  
Conc: 1.09 ng/ml




#18 AR-1242-3

R.T.: 5.027 min  
Delta R.T.: -0.005 min  
Response: 250485  
Conc: 1.85 ng/ml



#18 AR-1242-3


R.T.: 4.352 min  
Delta R.T.: -0.001 min  
Response: 188721  
Conc: 1.72 ng/ml



#19 AR-1242-4


R.T.: 5.139 min  
Delta R.T.: 0.004 min  
Response: 385307  
Conc: 3.55 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#19 AR-1242-4

R.T.: 4.444 min  
Delta R.T.: -0.004 min  
Response: 244586  
Conc: 2.33 ng/ml




#20 AR-1242-5

R.T.: 5.935 min  
Delta R.T.: 0.000 min  
Response: 453554  
Conc: 4.01 ng/ml



#20 AR-1242-5


R.T.: 5.003 min  
Delta R.T.: -0.005 min  
Response: 1008855  
Conc: 7.47 ng/ml



#21 AR-1248-1


R.T.: 4.942 min  
Delta R.T.: -0.002 min  
Response: 257066  
Conc: 2.42 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#21 AR-1248-1

R.T.: 4.148 min  
Delta R.T.: -0.004 min  
Response: 414753  
Conc: 3.84 ng/ml




#22 AR-1248-2

R.T.: 5.240 min  
Delta R.T.: 0.000 min  
Response: 588328  
Conc: 3.79 ng/ml



#22 AR-1248-2


R.T.: 4.405 min  
Delta R.T.: 0.000 min  
Response: 336733  
Conc: 2.33 ng/ml



#23 AR-1248-3


R.T.: 0.000 min  
Exp R.T.: 5.455 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#23 AR-1248-3

R.T.: 4.444 min  
Delta R.T.: -0.004 min  
Response: 244586  
Conc: 1.58 ng/ml




#24 AR-1248-4

R.T.: 5.891 min  
Delta R.T.: -0.004 min  
Response: 197150  
Conc: 1.04 ng/ml



#24 AR-1248-4


R.T.: 4.622 min  
Delta R.T.: -0.006 min  
Response: 426847  
Conc: 2.28 ng/ml



#25 AR-1248-5


R.T.: 5.935 min  
Delta R.T.: 0.000 min  
Response: 453554  
Conc: 2.38 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#25 AR-1248-5

R.T.: 5.029 min  
Delta R.T.: -0.021 min  
Response: 656603  
Conc: 3.53 ng/ml




#26 AR-1254-1

R.T.: 5.866 min  
Delta R.T.: 0.000 min  
Response: 1414772  
Conc: 6.91 ng/ml



#26 AR-1254-1


R.T.: 5.003 min  
Delta R.T.: -0.004 min  
Response: 1008855  
Conc: 3.46 ng/ml



#27 AR-1254-2


R.T.: 6.107 min  
Delta R.T.: 0.000 min  
Response: 1260835  
Conc: 4.05 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#27 AR-1254-2

R.T.: 5.165 min  
Delta R.T.: -0.005 min  
Response: 1135326  
Conc: 4.48 ng/ml




#28 AR-1254-3

R.T.: 6.493 min  
Delta R.T.: -0.010 min  
Response: 1602025  
Conc: 5.01 ng/ml



#28 AR-1254-3


R.T.: 5.616 min  
Delta R.T.: 0.015 min  
Response: 3991696  
Conc: 9.76 ng/ml



#29 AR-1254-4


R.T.: 6.812 min  
Delta R.T.: -0.008 min  
Response: 1605509  
Conc: 7.01 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#29 AR-1254-4

R.T.: 5.848 min  
Delta R.T.: -0.004 min  
Response: 553758  
Conc: 2.16 ng/ml




#30 AR-1254-5

R.T.: 7.277 min  
Delta R.T.: -0.004 min  
Response: 5523446  
Conc: 22.22 ng/ml



#30 AR-1254-5


R.T.: 6.298 min  
Delta R.T.: -0.008 min  
Response: 4777342  
Conc: 13.03 ng/ml



#31 AR-1260-1


R.T.: 6.685 min  
Delta R.T.: 0.000 min  
Response: 6393769  
Conc: 25.43 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#31 AR-1260-1

R.T.: 5.739 min  
Delta R.T.: -0.005 min  
Response: 5711993  
Conc: 20.01 ng/ml




#32 AR-1260-2

R.T.: 6.972 min  
Delta R.T.: -0.002 min  
Response: 6426918  
Conc: 22.36 ng/ml



#32 AR-1260-2


R.T.: 5.947 min  
Delta R.T.: -0.005 min  
Response: 6372995  
Conc: 18.38 ng/ml



#33 AR-1260-3


R.T.: 7.368 min  
Delta R.T.: -0.001 min  
Response: 8106399  
Conc: 38.34 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#33 AR-1260-3

R.T.: 6.111 min  
Delta R.T.: -0.003 min  
Response: 10543686  
Conc: 32.45 ng/ml




#34 AR-1260-4

R.T.: 7.612 min  
Delta R.T.: -0.002 min  
Response: 8084991  
Conc: 34.26 ng/ml



#34 AR-1260-4


R.T.: 6.623 min  
Delta R.T.: -0.006 min  
Response: 7370674  
Conc: 27.90 ng/ml



#35 AR-1260-5


R.T.: 7.957 min  
Delta R.T.: -0.002 min  
Response: 12429406  
Conc: 28.27 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#35 AR-1260-5

R.T.: 6.892 min  
Delta R.T.: -0.005 min  
Response: 14747724  
Conc: 23.91 ng/ml




#36 AR-1262-1

R.T.: 7.368 min  
Delta R.T.: -0.002 min  
Response: 8106399  
Conc: 23.95 ng/ml



#36 AR-1262-1


R.T.: 6.401 min  
Delta R.T.: -0.005 min  
Response: 3456506  
Conc: 19.27 ng/ml



#37 AR-1262-2


R.T.: 7.957 min  
Delta R.T.: -0.003 min  
Response: 12429406  
Conc: 21.97 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#37 AR-1262-2

R.T.: 6.623 min  
Delta R.T.: -0.006 min  
Response: 7370674  
Conc: 18.94 ng/ml




#38 AR-1262-3

R.T.: 8.272 min  
Delta R.T.: -0.004 min  
Response: 8260976  
Conc: 21.71 ng/ml



#38 AR-1262-3


R.T.: 7.202 min  
Delta R.T.: -0.006 min  
Response: 6525594  
Conc: 23.02 ng/ml



#39 AR-1262-4


R.T.: 8.359 min  
Delta R.T.: -0.003 min  
Response: 6151393  
Conc: 21.00 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#39 AR-1262-4

R.T.: 7.271 min  
Delta R.T.: -0.005 min  
Response: 11561781  
Conc: 20.08 ng/ml




#40 AR-1262-5

R.T.: 9.009 min  
Delta R.T.: -0.005 min  
Response: 5484376  
Conc: 27.37 ng/ml



#40 AR-1262-5


R.T.: 7.816 min  
Delta R.T.: -0.006 min  
Response: 5991192  
Conc: 23.02 ng/ml



#41 AR-1268-1


R.T.: 8.272 min  
Delta R.T.: 0.000 min  
Response: 8260976  
Conc: 13.63 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#41 AR-1268-1

R.T.: 7.202 min  
Delta R.T.: -0.006 min  
Response: 6525594  
Conc: 8.23 ng/ml




#42 AR-1268-2

R.T.: 8.359 min  
Delta R.T.: -0.008 min  
Response: 6151393  
Conc: 11.24 ng/ml



#42 AR-1268-2


R.T.: 7.271 min  
Delta R.T.: -0.007 min  
Response: 11561781  
Conc: 15.53 ng/ml



#43 AR-1268-3


R.T.: 8.584 min  
Delta R.T.: -0.008 min  
Response: 1724179  
Conc: 3.63 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024




#43 AR-1268-3

R.T.: 7.498 min  
Delta R.T.: -0.006 min  
Response: 4045754  
Conc: 6.39 ng/ml




#44 AR-1268-4

R.T.: 9.009 min  
Delta R.T.: -0.006 min  
Response: 5484376  
Conc: 26.95 ng/ml



#44 AR-1268-4


R.T.: 7.816 min  
Delta R.T.: -0.007 min  
Response: 5991192  
Conc: 22.84 ng/ml



#45 AR-1268-5

R.T.: 9.412 min  
Delta R.T.: -0.008 min  
Response: 2090862  
Conc: 1.43 ng/ml

Instrument :  
ECD\_R  
ClientSampleId :  
LOD-MDL-WATER-01-QT1-2024



#45 AR-1268-5

R.T.: 8.115 min  
Delta R.T.: -0.009 min  
Response: 2811182  
Conc: 1.47 ng/ml