

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012725\
 Data File : PR070070.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jan 2025 11:10
 Operator : AJ\MA
 Sample : Q1148-01DL 3X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 XA170DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/28/2025
 Supervised By :Ankita Jodhani 01/28/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 27 23:25:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011625CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jan 18 01:55:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	3.616	2.959	22366115	36796773	20.037m	5.788 #
2) SA Decachlor...	9.514	8.229	6618171	39791712	12.790	11.322

Target Compounds

16) L4 AR-1242-1	4.846	4.076	247.3E6	1281.4E6	10379.606	8454.440
17) L4 AR-1242-2	4.868	4.095	402.7E6	1953.8E6	10876.818	8868.131
18) L4 AR-1242-3	4.932	4.274	225.9E6	1011.5E6	9636.642	8687.630
19) L4 AR-1242-4	5.034	4.367	184.9E6	859.6E6	10433.660	7294.809 #
20) L4 AR-1242-5	5.822	4.916	196.4E6	1096.5E6	11157.850	8703.070

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012725\
Data File : PR070070.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jan 2025 11:10
Operator : AJ\MA
Sample : Q1148-01DL 3X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
XA170DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/28/2025
Supervised By :Ankita Jodhani 01/28/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 27 23:25:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011625CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 18 01:55:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

