

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012823\
 Data File : PR059229.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2023 02:15
 Operator : YP\AJ
 Sample : 01275-03
 Misc :
 ALS Vial : 76 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 EXGY6

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 10:19:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 02/02/2023
 Supervised By :Ankita Jodhani 02/02/2023

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)

R.T.	Response	Conc
5.85	67662724	716.31
6.09	94997083	650.70
6.48	83753929	542.75
6.79	56383304	487.50
7.25	71838971	580.03

(26) AR-1254-1 #2 (L6)

R.T.	Response	Conc
4.85	183701251	815.46
5.01	86432536	449.38
5.43	169460065	539.44
5.67	98398280	507.96
6.12	171209188	635.19

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012823\
 Data File : PR059229.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2023 02:15
 Operator : YP\AJ
 Sample : 01275-03
 Misc :
 ALS Vial : 76 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 EXGY6

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 10:19:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 02/02/2023
 Supervised By :Ankita Jodhani 02/02/2023

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
 R.T. Response Conc
 5.84 55526835 587.84
 6.09 94997083 650.70
 6.48 83753929 542.75
 6.79 56383304 487.50
 7.25 71838971 580.03

(26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 4.85 183701251 815.46
 5.01 86432536 449.38
 5.43 139376336 443.67
 5.67 98398280 507.96
 6.12 171209188 635.19


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012823\
Data File : PR059229.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jan 2023 02:15
Operator : YP\AJ
Sample : 01275-03
Misc :
ALS Vial : 76 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EXGY6

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 10:19:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 27 08:44:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
R.T. Response Conc
6.66 37837022 323.13
6.93 118886775 856.74
7.34 28808812 278.98
7.58 40136634 342.64
7.92 49290017 216.90

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.57 114320871 535.02
5.77 82565978 321.08
5.93 61304402 260.17
6.43 47612839 246.41
6.70 96939308 216.22

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012823\
 Data File : PR059229.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jan 2023 02:15
 Operator : YP\AJ
 Sample : 01275-03
 Misc :
 ALS Vial : 76 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 EXGY6

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 10:19:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 02/02/2023
 Supervised By :Ankita Jodhani 02/02/2023

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
 R.T. Response Conc
 6.66 37837022 323.13
 6.93 118886775 856.74
 7.34 28808812 278.98
 7.58 40136634 342.64
 7.92 49290017 216.90

(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 5.57 114320871 535.02
 5.77 82565978 321.08
 5.93 65140405 276.45
 6.43 47612839 246.41
 6.70 112195767 250.25