

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012919\
Data File : PR035299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2019 11:33
Operator : SM\SJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1660349

Manual Integrations
APPROVED

Sohil
1/30/2019 8:37:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 02:29:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR010419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 05 01:47:32 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.430min 19.479 ng/ml

response 23489440

(1) Tetrachloro-m-xylene #2 (SA)

3.506min 22.566 ng/ml

response 48066924

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012919\
Data File : PR035299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2019 11:33
Operator : SM\SJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1660349

Manual Integrations
APPROVED

Sohil
1/30/2019 8:37:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 02:29:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR010419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 05 01:47:32 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.430min 19.479 ng/ml

response 23489440

(1) Tetrachloro-m-xylene #2 (SA)

3.505min 19.883 ng/ml m

response 42351063

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012919\
Data File : PR035299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2019 11:33
Operator : SM\SJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1660349

Manual Integrations
APPROVED

Sohil
1/30/2019 8:37:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 02:29:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR010419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 05 01:47:32 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
R.T. Response Conc
7.19 57327686 371.73
7.44 78378275 375.01
7.72 97386187 412.63
8.02 62443506 374.01
8.34 150308491 382.59

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
6.01 134086940 374.21
6.20 179992706 374.03
6.35 162450546 376.86
6.81 138304092 376.86
7.05 374432624 366.07

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012919\
Data File : PR035299.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2019 11:33
Operator : SM\SJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1660349

Manual Integrations
APPROVED

Sohil
1/30/2019 8:37:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 30 02:29:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR010419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jan 05 01:47:32 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 #2 (L7)
R.T. Response Conc
7.19 57327686 371.73
7.44 78378275 375.01
7.72 87389050 370.27
8.02 62443506 374.01
8.34 150308491 382.59

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
6.01 134086940 374.21
6.20 179992706 374.03
6.35 162450546 376.86
6.81 138304092 376.86
7.05 374432624 366.07