

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012924\
 Data File : PR065303.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jan 2024 12:11
 Operator : AJ\MA
 Sample : P1157-02DL2 20X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AA9DL2

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 01/29/2024
 Supervised By :Sohil Jodhani 01/29/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 13:08:47 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 19 21:16:50 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.699	3.002	6195364	5122373	0.948	0.880
2)	SA Decachlor...	9.743	8.374	6101293	6490936	2.508	2.300m
Target Compounds							
16)	L4 AR-1242-1	4.942	4.148	73139352	75972282	479.227	487.899
17)	L4 AR-1242-2	4.964	4.166	203.5E6	193.0E6	908.239	909.813
18)	L4 AR-1242-3	5.031	4.350	44503563	95166930	312.022	822.335 #
19)	L4 AR-1242-4	5.133	4.442	93198703	95556470	818.960	853.136
20)	L4 AR-1242-5	5.934	5.002	126.5E6	232.4E6	1039.020	1627.335 #
26)	L6 AR-1254-1	5.863	5.002	168.8E6	232.4E6	771.471	751.362
27)	L6 AR-1254-2	6.104	5.164	252.2E6	173.3E6	766.762	647.851
28)	L6 AR-1254-3	6.502	5.596	224.0E6	295.2E6	669.735	684.926
29)	L6 AR-1254-4	6.819	5.846	155.3E6	172.4E6	678.546m	641.268
30)	L6 AR-1254-5	7.280	6.299	205.4E6	289.9E6	788.743	754.313

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012924\
Data File : PR065303.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2024 12:11
Operator : AJ\MA
Sample : P1157-02DL2 20X
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :

ECD_R

ClientSampleId :

C0AA9DL2

Manual Integrations

APPROVED

Reviewed By :Ankita Jodhani 01/29/2024

Supervised By :Sohil Jodhani 01/29/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 13:08:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 19 21:16:50 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

