

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012924\
 Data File : PR065324.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jan 2024 19:04
 Operator : AJ\MA
 Sample : AR1242CCC400
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12423353

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jan 29 20:55:50 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 19 21:16:50 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.700	3.002	122.5E6	109.3E6	18.756	18.776
2)	SA Decachlor...	9.740	8.373	94403780	108.1E6	38.800	38.283
Target Compounds							
16)	L4 AR-1242-1	4.942	4.147	57249601	58654411	375.114	376.682
17)	L4 AR-1242-2	4.963	4.166	87215542	81815993	389.217	385.721
18)	L4 AR-1242-3	5.028	4.349	56562367	44106848	396.568	381.126
19)	L4 AR-1242-4	5.132	4.443	45082726	41911781	396.153	374.192
20)	L4 AR-1242-5	5.933	5.002	46883354	55013079	384.947	385.212

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR012924\
Data File : PR065324.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jan 2024 19:04
Operator : AJ\MA
Sample : AR1242CCC400
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12423353

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jan 29 20:55:50 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR011224CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 19 21:16:50 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

