

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR013123\
 Data File : PR059375.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2023 07:36
 Operator : YP\AJ
 Sample : 01362-12 20X
 Misc :
 ALS Vial : 79 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 A4488

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 01 10:10:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.680	2.909	1233.1E6	7707782	454.552	1.938 #
2) SA Decachlor...	9.667	8.145	7139443	9257688	3.414	2.901
Target Compounds						
16) L4 AR-1242-1	4.928	4.021	6634.1E6	14095.3E6	89967.172	122683.831 #
17) L4 AR-1242-2	4.951	4.038	9992.8E6	20482.6E6	98846.488	128285.799 #
18) L4 AR-1242-3	5.015	4.215	3486.2E6	8846.5E6	53418.607	105046.781 #
19) L4 AR-1242-4	5.119	4.308	4133.9E6	8085.7E6	77485.460	101482.400 #
20) L4 AR-1242-5	5.913	4.850	1379.4E6	4557.2E6	25660.391	44086.935 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR013123\
Data File : PR059375.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2023 07:36
Operator : YP\AJ
Sample : 01362-12 20X
Misc :
ALS Vial : 79 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
A4488

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 01 10:10:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 27 08:44:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

