

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR020119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Feb 01 23:57:15 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035330.D 1600 =PR035332.D 800 =PR035331.D
 200 =PR035329.D 100 =PR035328.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.433	1.371	1.434	1.378	1.466	1.416	E6	2.86
2)	SA	Decachlorobiphenyl	2.000	1.980	2.037	2.130	2.129	2.055	E6	3.46
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1	3.930	3.400	3.728	3.912	4.439	3.882	E4	9.72
22)	L5	AR-1248-2	5.649	4.900	5.392	5.704	6.679	5.665	E4	11.47
23)	L5	AR-1248-3	6.470	5.711	6.226	6.426	7.190	6.405	E4	8.32
24)	L5	AR-1248-4	7.935	7.053	7.634	7.843	8.857	7.864	E4	8.30
25)	L5	AR-1248-5	7.498	6.684	7.288	7.364	8.341	7.435	E4	8.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR035330.D 1600 =PR035332.D 800 =PR035331.D
 200 =PR035329.D 100 =PR035328.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	3.474	3.296	3.453	3.394	3.666	3.457 E6	3.94	
2) SA	Decachlorobiphenyl	5.528	5.511	5.641	5.860	5.752	5.658 E6	2.63	
3) L1	AR-1016-1						0.000	-1.00	
4) L1	AR-1016-2						0.000	-1.00	
5) L1	AR-1016-3						0.000	-1.00	

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR020119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Feb 01 23:57:15 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035330.D 1600 =PR035332.D 800 =PR035331.D
 200 =PR035329.D 100 =PR035328.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	0.913	0.805	0.870	0.926	1.074	0.917 E5	10.81
22) L5	AR-1248-2	1.293	1.123	1.227	1.301	1.478	1.284 E5	10.09
23) L5	AR-1248-3	1.338	1.167	1.276	1.332	1.517	1.326 E5	9.56
24) L5	AR-1248-4	1.680	1.493	1.617	1.677	1.890	1.671 E5	8.60
25) L5	AR-1248-5	1.793	1.593	1.719	1.763	1.972	1.768 E5	7.77
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range