

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR020119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Feb 01 23:38:53 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035335.D 1600 =PR035337.D 800 =PR035336.D
 200 =PR035334.D 100 =PR035333.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.480	1.515	1.655	1.620	1.538	1.562	E6	4.70
2) SA	Decachlorobiphenyl	2.159	2.008	2.006	2.149	2.124	2.089	E6	3.66
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	1.074	0.985	1.094	1.188	1.209	1.110	E5	8.18
27) L6	AR-1254-2	1.731	1.601	1.770	1.886	1.893	1.776	E5	6.82
28) L6	AR-1254-3	1.878	1.770	1.929	1.996	1.967	1.908	E5	4.66
29) L6	AR-1254-4	1.583	1.494	1.620	1.685	1.671	1.610	E5	4.78
30) L6	AR-1254-5	1.633	1.554	1.668	1.727	1.668	1.650	E5	3.85
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR035335.D 1600 =PR035337.D 800 =PR035336.D
 200 =PR035334.D 100 =PR035333.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.580	3.652	3.987	3.978	3.782	3.796	E6	4.88
2) SA	Decachlorobiphenyl	5.987	5.575	5.530	5.996	5.925	5.803	E6	3.97
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR020119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Feb 01 23:38:53 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035335.D 1600 =PR035337.D 800 =PR035336.D
 200 =PR035334.D 100 =PR035333.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	3.744	3.526	3.888	4.099	4.071	3.866 E5	6.17
27) L6	AR-1254-2	3.296	3.087	3.401	3.606	3.603	3.399 E5	6.45
28) L6	AR-1254-3	5.728	5.426	5.927	6.146	6.053	5.856 E5	4.90
29) L6	AR-1254-4	3.829	3.620	3.945	4.153	4.077	3.925 E5	5.37
30) L6	AR-1254-5	5.444	5.140	5.555	5.789	5.715	5.529 E5	4.63
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range