

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR020119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Feb 02 01:12:46 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035340.D 1600 =PR035342.D 800 =PR035341.D
 200 =PR035339.D 100 =PR035338.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.563	1.490	1.566	1.304	1.795	1.544	E6	11.42
2)	SA	Decachlorobiphenyl	2.056	2.026	1.972	2.119	2.001	2.035	E6	2.77
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1	1.870	1.754	1.890	1.711	2.033	1.851	E5	6.84
37)	L8	AR-1262-2	3.219	3.156	3.326	2.994	3.337	3.206	E5	4.39
38)	L8	AR-1262-3	2.053	1.991	2.089	1.950	2.114	2.039	E5	3.34
39)	L8	AR-1262-4	1.518	1.467	1.537	1.447	1.548	1.503	E5	2.97
40)	L8	AR-1262-5	1.021	1.007	1.008	1.010	0.979	1.005	E5	1.55
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR035340.D 1600 =PR035342.D 800 =PR035341.D
 200 =PR035339.D 100 =PR035338.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	3.757	3.559	3.750	3.218	4.389	3.735 E6	11.41	
2) SA	Decachlorobiphenyl	5.649	5.559	5.482	5.811	5.594	5.619 E6	2.19	
3) L1	AR-1016-1						0.000	-1.00	
4) L1	AR-1016-2						0.000	-1.00	
5) L1	AR-1016-3						0.000	-1.00	

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR020119CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Feb 02 01:12:46 2019
 Response Via : Initial Calibration

Calibration Files

400 =PR035340.D 1600 =PR035342.D 800 =PR035341.D
 200 =PR035339.D 100 =PR035338.D

Compound			400	1600	800	200	100	Avg	%RSD
6)	L1	AR-1016-4						0.000	-1.00
7)	L1	AR-1016-5						0.000	-1.00
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1						0.000	-1.00
32)	L7	AR-1260-2						0.000	-1.00
33)	L7	AR-1260-3						0.000	-1.00
34)	L7	AR-1260-4						0.000	-1.00
35)	L7	AR-1260-5						0.000	-1.00
36)	L8	AR-1262-1	4.775	4.484	4.816	4.424	5.297	4.759 E5	7.28
37)	L8	AR-1262-2	8.943	8.588	9.136	8.289	9.300	8.851 E5	4.64
38)	L8	AR-1262-3	3.744	3.573	3.586	3.418	4.022	3.669 E5	6.24
39)	L8	AR-1262-4	6.299	6.124	6.380	5.971	6.550	6.265 E5	3.59
40)	L8	AR-1262-5	2.846	2.822	2.833	2.854	2.869	2.845 E5	0.64
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

 (#) = Out of Range