

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020119\
 Data File : PR035372.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Feb 2019 07:44
 Operator : SM\SJ
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1248303

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 04 00:20:58 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR020119CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Feb 02 01:45:00 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.426	3.501	36172856	85919889	23.327	28.011
2) SA Decachlor...	10.104	8.387	73885113	197.7E6	33.458	34.581
Target Compounds						
21) L5 AR-1248-1	5.590	4.560	18636005	43069601	480.082	469.478
22) L5 AR-1248-2	5.860	4.790	25972972	58534191	458.498	455.851
23) L5 AR-1248-3	6.064	4.830	29360313	59967642	458.419	452.277
24) L5 AR-1248-4	6.465	4.999	35535777	74843265	451.864	447.806
25) L5 AR-1248-5	6.504	5.382	33289534	76983273	447.754	435.424

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020119\
Data File : PR035372.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Feb 2019 07:44
Operator : SM\SJ
Sample : AR1248CCC400
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1248303

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 04 00:20:58 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR020119CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Feb 02 01:45:00 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

