

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020123\
 Data File : PR059422.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2023 22:07
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR16603309

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/02/2023
 Supervised By :Ankita Jodhani 02/02/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 00:53:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.693	2.908	54774853	81182343	20.192	20.417
2) SA Decachlor...	9.666	8.145	84747947	131.7E6	40.520	41.262
Target Compounds						
3) L1 AR-1016-1	4.927	4.019	34325528	55686640	396.284	401.075
4) L1 AR-1016-2	4.949	4.037	47040941	78259323	395.762	406.188
5) L1 AR-1016-3	5.014	4.215	31039100	41030784	405.822	404.161
6) L1 AR-1016-4	5.117	4.265	24700129	31161216	396.996	405.731
7) L1 AR-1016-5	5.436	4.481	24306798	36070530	395.435	351.009m
31) L7 AR-1260-1	6.659	5.567	44320652	85731089	378.503m	401.222m
32) L7 AR-1260-2	6.942	5.772	53348937	103.8E6	384.453m	403.564m
33) L7 AR-1260-3	7.335	5.930	41546250	94681464	402.321	401.813
34) L7 AR-1260-4	7.580	6.434	46197961	80963129	394.388	419.015
35) L7 AR-1260-5	7.920	6.700	89820643	184.2E6	395.254	410.920

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020123\
Data File : PR059422.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2023 22:07
Operator : YP\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR16603309

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/02/2023
Supervised By :Ankita Jodhani 02/02/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 00:53:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 27 08:44:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

