

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020123\
 Data File : PR059422.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2023 22:07
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 ECD_R
 LabSampleId :
 AR1660CCC400

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 02/02/2023
 Supervised By :Ankita Jodhani 02/02/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 00:53:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) AR-1016-1 (L1)
 R.T. Response Conc
 4.93 34325528 396.28
 4.95 47040941 395.76
 5.01 31039100 405.82
 5.12 24700129 397.00
 5.44 24306798 395.44

(3) AR-1016-1 #2 (L1)
 R.T. Response Conc
 4.02 55686640 401.08
 4.04 78259323 406.19
 4.21 41030784 404.16
 4.27 31161216 405.73
 4.48 42515632 413.73

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020123\
 Data File : PR059422.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2023 22:07
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1



Instrument :
 ECD_R
 LabSampleId :
 AR1660CCC400

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 02/02/2023
 Supervised By :Ankita Jodhani 02/02/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 00:53:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) AR-1016-1 #2 (L1)
 R.T. Response Conc
 4.93 34325528 396.28
 4.95 47040941 395.76
 5.01 31039100 405.82
 5.12 24700129 397.00
 5.44 24306798 395.44

(3) AR-1016-1 #2 (L1)
 R.T. Response Conc
 4.02 55686640 401.08
 4.04 78259323 406.19
 4.21 41030784 404.16
 4.27 31161216 405.73
 4.48 36070530 351.01

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020123\
 Data File : PR059422.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Feb 2023 22:07
 Operator : YP\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1



Instrument :
 ECD_R
 LabSampleId :
 AR1660CCC400

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 02 00:53:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jan 27 08:44:40 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 02/02/2023
 Supervised By :Ankita Jodhani 02/02/2023

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 (L7)

R.T.	Response	Conc
6.66	50540611	431.62
6.94	57054329	411.16
7.34	41546250	402.32
7.58	46197961	394.39
7.92	89820643	395.25

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
5.57	91345932	427.50
5.77	107684290	418.76
5.93	94681464	401.81
6.43	80963129	419.01
6.70	184228643	410.92

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020123\
Data File : PR059422.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Feb 2023 22:07
Operator : YP\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_R
LabSampleId :
AR1660CCC400

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 02/02/2023
Supervised By :Ankita Jodhani 02/02/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 02 00:53:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR012723CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jan 27 08:44:40 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
6.66 44320652 378.50
6.94 53348937 384.45
7.34 41546250 402.32
7.58 46197961 394.39
7.92 89820643 395.25

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.57 85731089 401.22
5.77 103776854 403.56
5.93 94681464 401.81
6.43 80963129 419.01
6.70 184228643 410.92