

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020822\
 Data File : PR053402.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Feb 2022 14:53
 Operator : AJ\MA
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ICVPR020822AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 08 15:07:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR020822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 08 15:00:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.576	2.819	96473893	68730083	50.286	51.544
2) SA Decachlor...	9.396	7.969	83956756	78309361	46.647	44.641
Target Compounds						
26) L6 AR-1254-1	5.695	4.717	35830287	43690821	473.101	461.573
27) L6 AR-1254-2	5.929	4.875	55410046	38533570	463.508	460.696
28) L6 AR-1254-3	6.323	5.286	58402147	61142818	464.661	463.860
29) L6 AR-1254-4	6.632	5.529	44611328	38883610	463.569	465.440
30) L6 AR-1254-5	7.084	5.967	48099652	56574902	466.157	482.460

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020822\
Data File : PR053402.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Feb 2022 14:53
Operator : AJ\MA
Sample : AR1254ICV500
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ICVPR020822AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 08 15:07:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR020822.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 08 15:00:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

