

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020822\
 Data File : PR053404.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Feb 2022 15:25
 Operator : AJ\MA
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 09 06:17:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR020822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 08 15:00:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.575	2.818	96962365	68283603	50.541	51.209
2) SA Decachlor...	9.395	7.969	85277858	78494074	47.381	44.747
Target Compounds						
26) L6 AR-1254-1	5.695	4.717	35759112	43532991	472.161	459.906
27) L6 AR-1254-2	5.929	4.874	55477508	38537497	464.072	460.743
28) L6 AR-1254-3	6.321	5.285	58988111	61071598	469.323	463.320
29) L6 AR-1254-4	6.632	5.529	44869145	38869247	466.248	465.268
30) L6 AR-1254-5	7.084	5.965	48491374	56488643	469.954	481.725



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020822\
Data File : PR053404.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Feb 2022 15:25
Operator : AJ\MA
Sample : AR1254CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 06:17:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR020822.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 08 15:00:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.575 min
Delta R.T.: 0.000 min
Response: 96962365
Conc: 50.54 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.818 min
Delta R.T.: 0.000 min
Response: 68283603
Conc: 51.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.395 min
Delta R.T.: 0.000 min
Response: 85277858
Conc: 47.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.969 min
Delta R.T.: 0.000 min
Response: 78494074
Conc: 44.75 ng/ml



#26 AR-1254-1


R.T.: 5.695 min
Delta R.T.: 0.000 min
Response: 35759112
Conc: 472.16 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 4.717 min
Delta R.T.: -0.001 min
Response: 43532991
Conc: 459.91 ng/ml




#27 AR-1254-2

R.T.: 5.929 min
Delta R.T.: 0.000 min
Response: 55477508
Conc: 464.07 ng/ml



#27 AR-1254-2


R.T.: 4.874 min
Delta R.T.: -0.001 min
Response: 38537497
Conc: 460.74 ng/ml



#28 AR-1254-3


R.T.: 6.321 min
Delta R.T.: 0.000 min
Response: 58988111
Conc: 469.32 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.285 min
Delta R.T.: -0.002 min
Response: 61071598
Conc: 463.32 ng/ml




#29 AR-1254-4

R.T.: 6.632 min
Delta R.T.: 0.000 min
Response: 44869145
Conc: 466.25 ng/ml



#29 AR-1254-4


R.T.: 5.529 min
Delta R.T.: -0.002 min
Response: 38869247
Conc: 465.27 ng/ml



#30 AR-1254-5

R.T.: 7.084 min
Delta R.T.: 0.000 min
Response: 48491374
Conc: 469.95 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 5.965 min
Delta R.T.: -0.002 min
Response: 56488643
Conc: 481.72 ng/ml