

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020822\
 Data File : PR053448.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Feb 2022 04:55
 Operator : AJ\MA
 Sample : N1425-08
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 20522

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 02/09/2022
 Supervised By :Ankita Jodhani 02/09/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Feb 09 06:48:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR020822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Feb 08 15:00:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.578	2.824	18935979	14636243	9.870m	10.976m
2) SA Decachlor...	9.402	7.980	28471857	39532762	15.819m	22.536m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR020822\
Data File : PR053448.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Feb 2022 04:55
Operator : AJ\MA
Sample : N1425-08
Misc :
ALS Vial : 38 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
20522


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/09/2022
Supervised By :Ankita Jodhani 02/09/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Feb 09 06:48:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR020822.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Feb 08 15:00:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.578 min
Delta R.T.: 0.004 min
Response: 18935979
Conc: 9.87 ng/ml

Instrument :
ECD_R
ClientSampleId :
20522


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 02/09/2022
Supervised By :Ankita Jodhani 02/09/2022




#1 Tetrachloro-m-xylene

R.T.: 2.824 min
Delta R.T.: 0.005 min
Response: 14636243
Conc: 10.98 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.402 min
Delta R.T.: 0.008 min
Response: 28471857
Conc: 15.82 ng/ml m



#2 Decachlorobiphenyl

R.T.: 7.980 min
Delta R.T.: 0.011 min
Response: 39532762
Conc: 22.54 ng/ml m